

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB88637

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP56337
calibration std. phosphate 0.5 ppm	WP56336
calibration std. phosphate 0.3 ppm	WP56335
calibration std. 0 ppm	WP56332
phosphate ICV std.	WP56339
phosphate CCV std.	WP56338
calibration std. phosphate 0.1 ppm	WP56334
calibration std. phosphate 0.05 ppm	WP56333
Combined reagent	WP56341
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP56340

Intercept: -0.0004Slope: 0.6501Regression: 0.999914

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/11/2017	10:49
2	CAL2	0.05	1	50	50	0.033	07/11/2017	10:49
3	CAL3	0.10	1	50	50	0.065	07/11/2017	10:50
4	CAL4	0.30	1	50	50	0.189	07/11/2017	10:50
5	CAL5	0.50	1	50	50	0.329	07/11/2017	10:51
6	CAL6	1.00	1	50	50	0.649	07/11/2017	10:51



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB88637

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.316	0.487	07/11/2017	10:52
2	ICB		1	50	50	0.000	0.001	07/11/2017	10:52
3	CCV1	0.50	1	50	50	0.330	0.508	07/11/2017	10:53
4	CCB1		1	50	50	0.000	0.001	07/11/2017	10:53
5	PB100566BL		1	50	50	0.000	0.001	07/11/2017	10:54
6	PB100566BS	0.50	1	50	50	0.318	0.490	07/11/2017	10:54
7	I4115-01		1	50	50	0.151	0.233	07/11/2017	10:55
8	I4115-01DUP		1	50	50	0.148	0.228	07/11/2017	10:55
9	I4115-01MS	0.50	1	50	50	0.450	0.693	07/11/2017	10:56
10	I4115-01MSD	0.50	1	50	50	0.448	0.690	07/11/2017	10:56
11	CCV2	0.50	1	50	50	0.329	0.507	07/11/2017	10:57
12	CCB2		1	50	50	0.000	0.001	07/11/2017	10:57