

7471A


LB 88639

INSTRUMENT ID: CV2

Linear

 μ Abs.:

38912



A= 0.0000e+000

B= 2.5819e-004

C= 1.7346e-003

Rho= 0.9998637

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	-50	0.000	-50				
0.2	0.200	0.195	-0.005	748	0.0 %	748				
2.5	2.500	2.599	0.099	10059	0.0 %	10059				
5.0	5.000	4.908	-0.092	19003	0.0 %	19003				
7.5	7.500	7.461	-0.039	28889	0.0 %	28889	0			
10.0	10.000	10.049	0.049	38912	0.0 %	38912				

LB88639 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 11 Jul 2017 13:46:53					
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.0000	7471A	PPB	11 Jul 2017 14:03:22	S	Std
0.2 - 1	50.2	748	-	0.2000	7471A	PPB	11 Jul 2017 14:05:40	S	Std
2.5 - 1	52.5	10059	-	2.5000	7471A	PPB	11 Jul 2017 14:07:58	S	Std
5.0 - 1	55	19003	-	5.0000	7471A	PPB	11 Jul 2017 14:10:15	S	Std
7.5 - 1	57.5	28889	-	7.5000	7471A	PPB	11 Jul 2017 14:12:32	S	Std
10.0 - 1	510	38912	-	10.0000	7471A	PPB	11 Jul 2017 14:17:37	S	Std
ICV56 - 1	ICV56	14817	3.8274	-	7471A	PPB	11 Jul 2017 14:20:17	U	SMPL
ICB56 - 1	ICB56	-198	-0.0494	-	7471A	PPB	11 Jul 2017 14:22:33	U	SMPL
CCV85 - 1	CCV85	19480	5.0313	-	7471A	PPB	11 Jul 2017 14:24:49	U	SMPL
CCB85 - 1	CCB85	-185	-0.0460	-	7471A	PPB	11 Jul 2017 14:27:05	U	SMPL
CRA - 1	CRA	755	0.1967	-	7471A	PPB	11 Jul 2017 14:29:21	U	SMPL
HighStd - 1	HighStd	38876	10.0393	-	7471A	PPB	11 Jul 2017 14:31:37	U	SMPL
ChkStd - 1	ChkStd	26630	6.8774	-	7471A	PPB	11 Jul 2017 14:33:52	U	SMPL
PB100540BL - 1	PBS	-223	-0.0558	-	7471A	PPB	11 Jul 2017 14:36:11	U	SMPL
PB100540BS - 1	LCSS	13965	3.6074	-	7471A	PPB	11 Jul 2017 14:38:28	U	SMPL
I4071-01 - 1	HR-06-070617-A	1600	0.4148	-	7471A	PPB	11 Jul 2017 14:45:36	U	SMPL
I4071-01DUP - 1	HR-06-070617-ADUP	1522	0.3947	-	7471A	PPB	11 Jul 2017 14:47:53	U	SMPL
I4071-01MS - 1	HR-06-070617-AMS	15982	4.1282	-	7471A	PPB	11 Jul 2017 14:50:11	U	SMPL
I4071-01MSD - 1	HR-06-070617-AMSD	15952	4.1204	-	7471A	PPB	11 Jul 2017 14:52:29	U	SMPL
I4071-04 - 1	HR-06-070617-B	776	0.2021	-	7471A	PPB	11 Jul 2017 14:54:45	U	SMPL
CCV86 - 1	CCV86	19614	5.0659	-	7471A	PPB	11 Jul 2017 14:57:01	U	SMPL
CCB86 - 1	CCB86	-193	-0.0481	-	7471A	PPB	11 Jul 2017 14:59:17	U	SMPL
I4071-07 - 1	HR-06-070617-C	237	0.0629	-	7471A	PPB	11 Jul 2017 15:01:33	U	SMPL
I4078-01 - 1	VNJ-228	783	0.2039	-	7471A	PPB	11 Jul 2017 15:03:50	U	SMPL
I4078-05 - 1	VNJ-211	1253	0.3253	-	7471A	PPB	11 Jul 2017 15:06:06	U	SMPL
I4081-01 - 1	VNJ-202	368615	HIGH	-	7471A	PPB	11 Jul 2017 15:08:22	X	
I4082-03 - 1	BIN-6-COMP	14029	3.6239	-	7471A	PPB	11 Jul 2017 15:10:13	U	SMPL
I4084-01 - 1	TP11-MH2111-COMP	13640	3.5235	-	7471A	PPB	11 Jul 2017 15:12:37	U	SMPL
I4084-06 - 1	TP9-MH2108-COMP	269	0.0712	-	7471A	PPB	11 Jul 2017 15:14:55	U	SMPL
I4108-01 - 1	BU-01-070617-A	687	0.1791	-	7471A	PPB	11 Jul 2017 15:17:12	U	SMPL
I4108-04 - 1	BU-01-070617-B	507	0.1326	-	7471A	PPB	11 Jul 2017 15:19:30	U	SMPL
I4108-07 - 1	BU-01-070617-C	9531	2.4626	-	7471A	PPB	11 Jul 2017 15:21:46	U	SMPL
CCV87 - 1	CCV87	19768	5.1057	-	7471A	PPB	11 Jul 2017 15:24:02	U	SMPL
CCB87 - 1	CCB87	-197	-0.0491	-	7471A	PPB	11 Jul 2017 15:26:19	U	SMPL
I4119-01 - 1	EO-03-070717-A	2911	0.7533	-	7471A	PPB	11 Jul 2017 15:28:35	U	SMPL
I4119-04 - 1	EO-03-070717-B	1401	0.3635	-	7471A	PPB	11 Jul 2017 15:30:51	U	SMPL
I4119-07 - 1	EO-03-070717-C	1270	0.3296	-	7471A	PPB	11 Jul 2017 15:33:07	U	SMPL
PB100541BL - 1	PBS	-73	-0.0171	-	7471A	PPB	11 Jul 2017 15:35:24	U	SMPL
PB100541BS - 1	LCSS	14539	3.7556	-	7471A	PPB	11 Jul 2017 15:37:40	U	SMPL
I4101-01 - 1	EP-17	610	0.1592	-	7471A	PPB	11 Jul 2017 15:39:57	U	SMPL
I4101-02 - 1	EP-16	617	0.1610	-	7471A	PPB	11 Jul 2017 15:42:15	U	SMPL
I4101-03 - 1	EP-11	327	0.0862	-	7471A	PPB	11 Jul 2017 15:44:32	U	SMPL
I4101-03DUP - 1	EP-11DUP	303	0.0800	-	7471A	PPB	11 Jul 2017 15:46:50	U	SMPL
I4101-03MS - 1	EP-11MS	15239	3.9363	-	7471A	PPB	11 Jul 2017 15:49:06	U	SMPL
CCV88 - 1	CCV88	19116	4.9374	-	7471A	PPB	11 Jul 2017 15:51:22	U	SMPL
CCB88 - 1	CCB88	-179	-0.0445	-	7471A	PPB	11 Jul 2017 15:53:38	U	SMPL
I4101-03MSD - 1	EP-11MSD	15261	3.9420	-	7471A	PPB	11 Jul 2017 15:55:55	U	SMPL
I4101-04 - 1	EP-10	605	0.1579	-	7471A	PPB	11 Jul 2017 15:58:12	U	SMPL
I4105-06 - 1	CIY-SB-020304-0-6	19632	5.0706	-	7471A	PPB	11 Jul 2017 16:00:28	U	SMPL
I4113-01 - 1	SB-4(10.0-10.5)	29	0.0092	-	7471A	PPB	11 Jul 2017 16:02:45	U	SMPL
I4113-02 - 1	SB-4	369	0.0970	-	7471A	PPB	11 Jul 2017 16:05:02	U	SMPL
I4113-03 - 1	SB-5(8.0-8.5)	3301	0.8540	-	7471A	PPB	11 Jul 2017 16:07:19	U	SMPL
I4113-04 - 1	SB-5	20720	5.3515	-	7471A	PPB	11 Jul 2017 16:09:36	U	SMPL
I4113-05 - 1	SB-6(11.0-11.5)	1772	0.4593	-	7471A	PPB	11 Jul 2017 16:11:54	U	SMPL
I4113-06 - 1	SB-6	2458	0.6364	-	7471A	PPB	11 Jul 2017 16:14:12	U	SMPL
I4114-01 - 1	NAA-SAND	164	0.0441	-	7471A	PPB	11 Jul 2017 16:16:28	U	SMPL
CCV89 - 1	CCV89	19202	4.9596	-	7471A	PPB	11 Jul 2017 16:18:45	U	SMPL
CCB89 - 1	CCB89	-192	-0.0478	-	7471A	PPB	11 Jul 2017 16:21:01	U	SMPL
I4118-01 - 1	NAA-PIPE-SAND	262	0.0694	-	7471A	PPB	11 Jul 2017 16:23:18	U	SMPL
I4120-01 - 1	TP-12-COMP	1092	0.2837	-	7471A	PPB	11 Jul 2017 16:25:34	U	SMPL
I4127-01 - 1	SB-7	7378	1.9067	-	7471A	PPB	11 Jul 2017 16:27:50	U	SMPL
I4127-02 - 1	SB-7	571	0.1492	-	7471A	PPB	11 Jul 2017 16:30:07	U	SMPL
I4127-03 - 1	SB-8	266	0.0704	-	7471A	PPB	11 Jul 2017 16:32:23	U	SMPL
I4127-04 - 1	SB-8	234	0.0622	-	7471A	PPB	11 Jul 2017 16:34:40	U	SMPL
I4132-01 - 1	CLINTON-AVE-69KV-COMP-(DRUMS-1-6)	7450	1.9253	-	7471A	PPB	11 Jul 2017 16:36:57	U	SMPL
PB100542BL - 1	PBS	110	0.0301	-	7471A	PPB	11 Jul 2017 16:39:15	U	SMPL
PB100542BS - 1	LCSS	13886	3.5870	-	7471A	PPB	11 Jul 2017 16:41:33	U	SMPL
I4130-01 - 1	E2-P10-BASE	34356	8.8722	-	7471A	PPB	11 Jul 2017 16:43:49	U	SMPL
CCV90 - 1	CCV90	18674	4.8232	-	7471A	PPB	11 Jul 2017 16:46:06	U	SMPL
CCB90 - 1	CCB90	-215	-0.0538	-	7471A	PPB	11 Jul 2017 16:48:24	U	SMPL
I4130-02 - 1	E2-P9-BASE	616258	HIGH	-	7471A	PPB	11 Jul 2017 16:50:41	X	
I4130-03 - 1	E2-P6-WSW	36154	9.3365	-	7471A	PPB	11 Jul 2017 16:52:31	U	SMPL
I4130-04 - 1	E2-P7-BASE	302112	HIGH	-	7471A	PPB	11 Jul 2017 16:55:18	X	
I4130-05 - 1	E2-P8-BASE	293217	HIGH	-	7471A	PPB	11 Jul 2017 16:57:11	X	
I4130-06 - 1	E2-HS1-SSW	26221	6.7718	-	7471A	PPB	11 Jul 2017 16:59:10	U	SMPL
I4130-07 - 1	E2-HS1-NSW	1037	0.2695	-	7471A	PPB	11 Jul 2017 17:01:33	U	SMPL
I4130-08 - 1	E2-HS1-WSW	26483	6.8395	-	7471A	PPB	11 Jul 2017 17:03:52	U	SMPL
I4130-09 - 1	E2-HS1-BASE	3754	0.9710	-	7471A	PPB	11 Jul 2017 17:06:09	U	SMPL
I4130-10 - 1	E2-HS1-ESW	15262	3.9423	-	7471A	PPB	11 Jul 2017 17:08:28	U	SMPL
I4130-11 - 1	E2-HS2-WSW	5719	1.4783	-	7471A	PPB	11 Jul 2017 17:10:45	U	SMPL
CCV91 - 1	CCV91	19796	5.1129	-	7471A	PPB	11 Jul 2017 17:13:02	U	SMPL
CCB91 - 1	CCB91	-23	-0.0042	-	7471A	PPB	11 Jul 2017 17:15:19	U	SMPL
I4130-12 - 1	E2-HS2-SSW	28281	7.3037	-	7471A	PPB	11 Jul 2017 17:17:37	U	SMPL
I4130-13 - 1	E2-HS2-NSW	6596	1.7048	-	7471A	PPB	11 Jul 2017 17:19:54	U	SMPL
I4130-14 - 1	E2-HS2-ESW	123396	31.8617	-	7471A	PPB	11 Jul 2017 17:22:11	U	SMPL
I4130-15 - 1	E2-HS2-BASE	1849	0.4791	-	7471A	PPB	11 Jul 2017 17:24:28	U	SMPL
I4130-16 - 1	E2-HS3-NSW	37909	9.7896	-	7471A	PPB	11 Jul 2017 17:26:49	U	SMPL
I4130-17 - 1	E2-HS3-ESW	35404	9.1428	-	7471A	PPB	11 Jul 2017 17:29:06	U	SMPL
I4130-18 - 1	E2-HS3-WSW	43338	11.1913	-	7471A	PPB	11 Jul 2017 17:31:25	U	SMPL
I4130-19 - 1	E2-HS3-SSW	31673	8.1795	-	7471A	PPB	11 Jul 2017 17:33:44	U	SMPL
I4130-20 - 1	E2-HS3-BASE	746	0.1943	-	7471A	PPB	11 Jul 2017 17:36:04	U	SMPL
I4130-20DUP - 1	E2-HS3-BASEDUP	644	0.1680	-	7471A	PPB	11 Jul 2017 17:43:40	U	SMPL
CCV92 - 1	CCV92	18946	4.8935	-	7471A	PPB	11 Jul 2017 17:45:57	U	SMPL

LB88639 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 11 Jul 2017 13:46:53					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB92 - 1	CCB92	-23	-0.0042		-7471A	PPB	11 Jul 2017 17:48:14	U	SMPL
I4130-20MS - 1	E2-HS3-BASEMS	15429	3.9854		-7471A	PPB	11 Jul 2017 17:50:32	U	SMPL
I4130-20MSD - 1	E2-HS3-BASEMSD	15267	3.9436		-7471A	PPB	11 Jul 2017 17:52:49	U	SMPL
I4081-01DLX10 - 1		51476	13.2925		-7471A	PPB	11 Jul 2017 17:55:06	U	SMPL
I4071-01LX5 - 1		4	0.0028		-7471A	PPB	11 Jul 2017 17:57:23	U	SMPL
I4071-01A - 1		16927	4.3722		-7471A	PPB	11 Jul 2017 17:59:42	U	SMPL
I4101-03LX5 - 1		21	0.0072		-7471A	PPB	11 Jul 2017 18:01:59	U	SMPL
I4101-03A - 1		20626	5.3272		-7471A	PPB	11 Jul 2017 18:04:16	U	SMPL
I4130-20LX5 - 1		94	0.0260		-7471A	PPB	11 Jul 2017 18:06:34	U	SMPL
I4130-20A - 1		16677	4.3076		-7471A	PPB	11 Jul 2017 18:08:51	U	SMPL
I4081-01DL2X50 - 1		10447	2.6991		-7471A	PPB	11 Jul 2017 18:11:08	U	SMPL
CCV93 - 1	CCV93	19439	5.0208		-7471A	PPB	11 Jul 2017 18:13:25	U	SMPL
CCB93 - 1	CCB93	-200	-0.0499		-7471A	PPB	11 Jul 2017 18:15:42	U	SMPL
I4130-02DLX10 - 1		107211	27.6829		-7471A	PPB	11 Jul 2017 18:18:00	U	SMPL
I4130-04DLX10 - 1		39840	10.2882		-7471A	PPB	11 Jul 2017 18:20:17	U	SMPL
I4130-05DLX10 - 1		39001	10.0715		-7471A	PPB	11 Jul 2017 18:22:36	U	SMPL
I4130-14DLX10 - 1		12201	3.1520		-7471A	PPB	11 Jul 2017 18:24:55	U	SMPL
I4130-02DL2X50 - 1		21919	5.6611		-7471A	PPB	11 Jul 2017 18:27:14	U	SMPL
I4130-04DL2X50 - 1		8077	2.0872		-7471A	PPB	11 Jul 2017 18:29:31	U	SMPL
I4130-05DL2X50 - 1		7513	1.9415		-7471A	PPB	11 Jul 2017 18:31:49	U	SMPL
I4130-18DLX2 - 1		23319	6.0225		-7471A	PPB	11 Jul 2017 18:34:06	U	SMPL
CCV94 - 1	CCV94	18745	4.8416		-7471A	PPB	11 Jul 2017 18:36:24	U	SMPL
CCB94 - 1	CCB94	-35	-0.0073		-7471A	PPB	11 Jul 2017 18:38:42	U	SMPL