



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88640

Temp.: 23.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/11/2017	14:30
2	CAL2	1	Water			7.01	07/11/2017	14:31
3	CAL3	1	Water			10.03	07/11/2017	14:32
4	ICV	1	Water			7.00	07/11/2017	14:35
5	CCV1	1	Water			2.01	07/11/2017	14:36
6	I4139-01	1	Water			4.42	07/11/2017	14:37
7	I4139-01DUP	1	Water			4.43	07/11/2017	14:38
8	CCV2	1	Water			12.01	07/11/2017	14:40

2B88640

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i4139

WorkList ID : 100576

Date : 7/11/2017 2:15:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4139-01-17	pH	Cool 4 deg C	STER11	D31	POOLWATER	07/03/2017	SM 4500-PH B

Date/Time 07-11-17 2:15 PM
 Received by: JP
 Relinquished by: CP

Date/Time 07-11-17 3:20 PM
 Received by: CP
 Relinquished by: JP