

Daily Analysis Runlog For Sequence/QC Batch ID # LB88642

Review By

Review On

SubDirectory

LB88642

Test :

Anions

STD. NAME		STD REF.#				
ICAL Standard		WP56304,WP56305,WP56306,WP56307,WP56308,WP56309,WP56310				
ICV Standard		WP56311				
CCV Standard		WP56306				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP56311				
Chk Standard		WP56312,WP56313				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	STD1	STD1	CAL1	07/11/17 10:41		OK
2	STD2	STD2	CAL2	07/11/17 11:10		OK
3	STD3	STD3	CAL3	07/11/17 11:39		OK
4	STD4	STD4	CAL4	07/11/17 12:08		OK
5	STD5	STD5	CAL5	07/11/17 12:37		OK
6	STD6	STD6	CAL6	07/11/17 13:06		OK
7	STD7	STD7	CAL7	07/11/17 13:35		OK
8	CCV1	CCV1	CCV	07/11/17 14:05		OK
9	CCB1	CCB1	CCB	07/11/17 14:34		OK
10	ICV1	ICV1	ICV	07/11/17 15:03		OK
11	ICB1	ICB1	ICB	07/11/17 15:32		OK
12	LB88642BLW	LB88642BLW	MB	07/11/17 16:01		OK
13	LB88642BSW	LB88642BSW	LCS	07/11/17 16:30		OK
14	I4139-01	POOLWATER	SAM	07/11/17 16:59		OK
15	I4139-01DUP	POOLWATERDUP	DUP	07/11/17 17:27		OK
16	I4139-01MS	POOLWATERMS	MS	07/11/17 17:56		OK
17	CCV2	CCV2	CCV	07/11/17 18:25		OK
18	CCB2	CCB2	CCB	07/11/17 18:54		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	7/3/2017 12:00:00 AM
Project:	LB88642	Date Received:	7/11/2017 12:00:00 AM
Client Sample ID:	POOLWATER	SDG No.:	LB88642
Lab Sample ID:	I4139-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Bromide	0	U	1	0.066		0.5	mg/L		07/11/2017	300.0
	Chloride	0	U	1	0.075		0.15	mg/L		07/11/2017	300.0
	Fluoride	0	U	1	0.043		0.1	mg/L		07/11/2017	300.0
	Nitrate	0	U	1	0.027		0.13	mg/L		07/11/2017	300.0
	Nitrite	0	U	1	0.022		0.15	mg/L		07/11/2017	300.0
	Phosphorus, Ortho	0	U	1	0.01		0.25	mg/L		07/11/2017	300.0
	Sulfate	0	U	1	0.13		0.75	mg/L		07/11/2017	300.0



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	7/3/2017 12:00:00 AM
Project:	LB88642	Date Received:	7/11/2017 12:00:00 AM
Client Sample ID:	POOLWATER	SDG No.:	LB88642
Lab Sample ID:	I4139-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB88642

SDG No.: LB88642

Contract: _____ Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Bromide	9.9	10	99	0	90 - 110		07/11/2017	14:05	LB88642
	Chloride	3	3	100	0	90 - 110		07/11/2017	14:05	LB88642
	Fluoride	2	2	100	0	90 - 110		07/11/2017	14:05	LB88642
	Phosphorus, Ortho	5.1	5	102	0	90 - 110		07/11/2017	14:05	LB88642
	Nitrite	2.9	3	96.7	0	90 - 110		07/11/2017	14:05	LB88642
	Nitrate	2.5	2.5	100	0	90 - 110		07/11/2017	14:05	LB88642
	Sulfate	14.3	15	95.3	0	90 - 110		07/11/2017	14:05	LB88642



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB88642

SDG No.: LB88642

Contract: _____ Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Bromide	10.3	10	103	0	90 - 110		07/11/2017	15:03	LB88642
	Chloride	3	3	100	0	90 - 110		07/11/2017	15:03	LB88642
	Fluoride	2.1	2	105	0	90 - 110		07/11/2017	15:03	LB88642
	Phosphorus, Ortho	5.3	5	106	0	90 - 110		07/11/2017	15:03	LB88642
	Nitrite	3	3	100	0	90 - 110		07/11/2017	15:03	LB88642
	Nitrate	2.5	2.5	100	0	90 - 110		07/11/2017	15:03	LB88642
	Sulfate	14.7	15	98	0	90 - 110		07/11/2017	15:03	LB88642

Sample ID	Analyte	Result	True Value	%	%	Acceptance	M	Analysis	Analysis	Run
		mg/L								
CCV2	Bromide	10	10	100	0	90 - 110		07/11/2017	18:25	LB88642
	Chloride	3	3	100	0	90 - 110		07/11/2017	18:25	LB88642
	Fluoride	2	2	100	0	90 - 110		07/11/2017	18:25	LB88642
	Phosphorus, Ortho	5.2	5	104	0	90 - 110		07/11/2017	18:25	LB88642
	Nitrite	2.9	3	96.7	0	90 - 110		07/11/2017	18:25	LB88642
	Nitrate	2.5	2.5	100	0	90 - 110		07/11/2017	18:25	LB88642
	Sulfate	14.4	15	96	0	90 - 110		07/11/2017	18:25	LB88642



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88642

SDG No.: LB88642

Contract: Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Bromide	0.066	+/-0.5	U		0.5		07/11/2017	14:34	LB88642
	Chloride	0.075	+/-0.15	U		0.15		07/11/2017	14:34	LB88642
	Fluoride	0.043	+/-0.1	U		0.1		07/11/2017	14:34	LB88642
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		07/11/2017	14:34	LB88642
	Nitrite	0.022	+/-0.15	U		0.15		07/11/2017	14:34	LB88642
	Nitrate	0.027	+/-0.13	U		0.13		07/11/2017	14:34	LB88642
	Sulfate	0.13	+/-0.75	U		0.75		07/11/2017	14:34	LB88642



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88642

SDG No.: LB88642

Contract:

Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Bromide	0.066	+/-0.5	U		0.5		07/11/2017	15:32	LB88642
	Chloride	0.075	+/-0.15	U		0.15		07/11/2017	15:32	LB88642
	Fluoride	0.043	+/-0.1	U		0.1		07/11/2017	15:32	LB88642
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		07/11/2017	15:32	LB88642
	Nitrite	0.022	+/-0.15	U		0.15		07/11/2017	15:32	LB88642
	Nitrate	0.027	+/-0.13	U		0.13		07/11/2017	15:32	LB88642
	Sulfate	0.13	+/-0.75	U		0.75		07/11/2017	15:32	LB88642



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88642

SDG No.: LB88642

Contract:

Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB2	Bromide	0.066	+/-0.5	U		0.5		07/11/2017	18:54	LB88642
	Chloride	0.075	+/-0.15	U		0.15		07/11/2017	18:54	LB88642
	Fluoride	0.043	+/-0.1	U		0.1		07/11/2017	18:54	LB88642
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		07/11/2017	18:54	LB88642
	Nitrite	0.022	+/-0.15	U		0.15		07/11/2017	18:54	LB88642
	Nitrate	0.027	+/-0.13	U		0.13		07/11/2017	18:54	LB88642
	Sulfate	0.13	+/-0.75	U		0.75		07/11/2017	18:54	LB88642



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88642

SDG No.: LB88642

Contract: Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB88642

SDG No.: LB88642

Instrument: IC-1

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB88642BLW		WATER		Batch Number:				Prep Date:		
	Bromide	0.066	<0.5	U		0.5		07/11/2017	16:01	LB88642
	Chloride	0.075	<0.15	U		0.15		07/11/2017	16:01	LB88642
	Fluoride	0.043	<0.1	U		0.1		07/11/2017	16:01	LB88642
	Phosphorus, Ortho	0.01	<0.25	U		0.25		07/11/2017	16:01	LB88642
	Nitrite	0.022	<0.15	U		0.15		07/11/2017	16:01	LB88642
	Nitrate	0.027	<0.13	U		0.13		07/11/2017	16:01	LB88642
	Sulfate	0.13	<0.75	U		0.75		07/11/2017	16:01	LB88642



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB88642 level: low sdg no.: LB88642
contract: _____ lab code: CHEM case no.: LB88642 sas no.: LB88642
matrix: WATER sample id: I4139-01 client id: POOLWATERMS
Percent Solids for Sample: 0 Spiked ID: I4139-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Bromide	mg/L	80 - 120	10.2		0.066	U	10	102		
Chloride	mg/L	80 - 120	3.1		0.075	U	3	103.3		
Fluoride	mg/L	80 - 120	2.2		0.043	U	2	110		
Phosphorus, Ortho	mg/L	80 - 120	5.5		0.01	U	5	110		
Nitrite	mg/L	80 - 120	3		0.022	U	3	100		
Nitrate	mg/L	80 - 120	2.6		0.027	U	2.5	104		
Sulfate	mg/L	80 - 120	15.4		0.13	U	15	102.7		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88642 **Level:** LOW **SDG No.:** LB88642
Contract: _____ **Lab Code:** CHEM **Case No.:** LB88642 **SAS No.:** LB88642
Matrix: WATER **Sample ID:** I4139-01 **Client ID:** POOLWATERDUP
Percent Solids for Sample: 0 **Duplicate ID** I4139-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Bromide	mg/L	20	0.066	U	0	U			
Chloride	mg/L	20	0.075	U	0	U			
Fluoride	mg/L	20	0.043	U	0	U			
Phosphorus, Ortho	mg/L	20	0.01	U	0	U			
Nitrite	mg/L	20	0.022	U	0	U			
Nitrate	mg/L	20	0.027	U	0	U			
Sulfate	mg/L	20	0.13	U	0	U			



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB88642

SDG No.: LB88642

Contract:

Lab Code: CHEM

Case No.: LB88642

SAS No.: LB88642

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB88642BSW							
Bromide	mg/L	10	10		100	90 - 110	
Chloride	mg/L	3	3		100	90 - 110	
Fluoride	mg/L	2	2.1		105	90 - 110	
Phosphorus, Ortho	mg/L	5	5.3		106	90 - 110	
Nitrite	mg/L	3	3		100	90 - 110	
Nitrate	mg/L	2.5	2.5		100	90 - 110	
Sulfate	mg/L	15	14.8		98.7	90 - 110	
