


7470A

L388655
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32337



A= 0.0000e+000

B= 3.0828e-004

C= 4.6583e-002

Rho= 0.9998122

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.061	0.061	48	0.000	48				
0.2	0.200	0.241	0.041	632	0.0 %	632				
2.5	2.500	2.417	-0.083	7688	0.0 %	7688				
5.0	5.000	4.889	-0.111	15709	0.0 %	15709				
7.5	7.500	7.576	0.076	24424	0.0 %	24424				
10.0	10.000	10.015	0.015	32337	0.0 %	32337				

LB88655 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jul 2017 10:07:58

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	48	-	0.0000	7470A	PPB	12 Jul 2017 10:39:33	S	Std
0.2 - 1	50.2	632	-	0.2000	7470A	PPB	12 Jul 2017 10:41:41	S	Std
2.5 - 1	52.5	7688	-	2.5000	7470A	PPB	12 Jul 2017 10:43:50	S	Std
5.0 - 1	55	15709	-	5.0000	7470A	PPB	12 Jul 2017 10:45:58	S	Std
7.5 - 1	57.5	24424	-	7.5000	7470A	PPB	12 Jul 2017 10:48:06	S	Std
10.0 - 1	510	32337	-	10.0000	7470A	PPB	12 Jul 2017 10:50:14	S	Std
ICV57 - 1	ICV57	13090	4.0819	-	7470A	PPB	12 Jul 2017 10:53:29	U	SMPL
ICB57 - 1	ICB57	53	0.0629	-	7470A	PPB	12 Jul 2017 10:55:36	U	SMPL
CCV95 - 1	CCV95	15744	4.9001	-	7470A	PPB	12 Jul 2017 10:57:43	U	SMPL
CCB95 - 1	CCB95	16	0.0515	-	7470A	PPB	12 Jul 2017 10:59:50	U	SMPL
CRA - 1	CRA	668	0.2525	-	7470A	PPB	12 Jul 2017 11:01:58	U	SMPL
HighStd - 1	HighStd	32613	10.1004	-	7470A	PPB	12 Jul 2017 11:04:04	U	SMPL
ChkStd - 1	ChkStd	22462	6.9711	-	7470A	PPB	12 Jul 2017 11:06:11	U	SMPL
PB100572BL - 1	PBW	-211	-0.0185	-	7470A	PPB	12 Jul 2017 11:08:20	U	SMPL
PB100572BS - 1	LCSW	11601	3.6229	-	7470A	PPB	12 Jul 2017 11:10:29	U	SMPL
PB100572TB - 1	PB100531TB	-95	0.0173	-	7470A	PPB	12 Jul 2017 11:12:37	U	SMPL
I4120-05 - 1	TP-12-COMP	-31	0.0370	-	7470A	PPB	12 Jul 2017 11:14:46	U	SMPL
I4127-05 - 1	SB-7	-57	0.0290	-	7470A	PPB	12 Jul 2017 11:16:54	U	SMPL
I4127-06 - 1	SB-8	-15	0.0420	-	7470A	PPB	12 Jul 2017 11:19:02	U	SMPL
I4132-03 - 1	CLINTON-AVE-69KV-COMP-(DRUMS-1-6)	175	0.1005	-	7470A	PPB	12 Jul 2017 11:21:09	U	SMPL
CCV96 - 1	CCV96	15914	4.9525	-	7470A	PPB	12 Jul 2017 11:23:16	U	SMPL
CCB96 - 1	CCB96	93	0.0753	-	7470A	PPB	12 Jul 2017 11:25:23	U	SMPL
I4132-03DUP - 1	CLINTON-AVE-69KV-COMP-(DRUMS-1-6)DUP	244	0.1218	-	7470A	PPB	12 Jul 2017 11:27:30	U	SMPL
I4132-03MS - 1	CLINTON-AVE-69KV-COMP-(DRUMS-1-6)MS	12672	3.9531	-	7470A	PPB	12 Jul 2017 11:29:37	U	SMPL
I4132-03MSD - 1	CLINTON-AVE-69KV-COMP-(DRUMS-1-6)MSD	12471	3.8911	-	7470A	PPB	12 Jul 2017 11:31:44	U	SMPL
I4132-03LX5 - 1		-54	0.0299	-	7470A	PPB	12 Jul 2017 11:33:51	U	SMPL
I4132-03A - 1		13547	4.2228	-	7470A	PPB	12 Jul 2017 11:35:59	U	SMPL
CCV97 - 1	CCV97	15870	4.9389	-	7470A	PPB	12 Jul 2017 11:38:06	U	SMPL
CCB97 - 1	CCB97	84	0.0725	-	7470A	PPB	12 Jul 2017 11:40:15	U	SMPL