


7471A

LB 88661
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38774



A= 0.0000e+000

B= 2.5955e-004

C= -3.8226e-002

Rho= 0.9998771

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	69	0.000	69				
0.2	0.200	0.204	0.004	933	0.0 %	933				
2.5	2.500	2.494	-0.006	9755	0.0 %	9755				
5.0	5.000	5.096	0.096	19780	0.0 %	19780				
7.5	7.500	7.401	-0.099	28662	0.0 %	28662	0			
10.0	10.000	10.026	0.026	38774	0.0 %	38774				

LB88661 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 12 Jul 2017 14:32:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	69	-	0.0000	7471A	PPB	12 Jul 2017 14:37:54	S	Std
0.2 - 1	50.2	933	-	0.2000	7471A	PPB	12 Jul 2017 14:40:11	S	Std
2.5 - 1	52.5	9755	-	2.5000	7471A	PPB	12 Jul 2017 14:42:29	S	Std
5.0 - 1	55	19780	-	5.0000	7471A	PPB	12 Jul 2017 14:44:46	S	Std
7.5 - 1	57.5	28662	-	7.5000	7471A	PPB	12 Jul 2017 14:47:03	S	Std
10.0 - 1	50	38774	-	10.0000	7471A	PPB	12 Jul 2017 14:49:21	S	Std
ICV58 - 1	ICV58	16187	4.1632	-7471A	PPB	12 Jul 2017 14:52:11	U	SMPL	
ICB58 - 1	ICB58	-19	-0.0432	-7471A	PPB	12 Jul 2017 14:54:27	U	SMPL	
CCV01 - 1	CCV01	18839	4.8515	-7471A	PPB	12 Jul 2017 14:56:42	U	SMPL	
CCB01 - 1	CCB01	-16	-0.0424	-7471A	PPB	12 Jul 2017 14:58:59	U	SMPL	
CRA - 1	CRA	929	0.2029	-7471A	PPB	12 Jul 2017 15:01:15	U	SMPL	
HighStd - 1	HighStd	38608	9.9827	-7471A	PPB	12 Jul 2017 15:03:31	U	SMPL	
ChkStd - 1	ChkStd	26912	6.9469	-7471A	PPB	12 Jul 2017 15:05:47	U	SMPL	
PB100574BL - 1	PBS	-163	-0.0805	-7471A	PPB	12 Jul 2017 15:08:05	U	SMPL	
PB100574BS - 1	LCSS	14247	3.6597	-7471A	PPB	12 Jul 2017 15:10:24	U	SMPL	
I4140-01 - 1	EP-03	521	0.0970	-7471A	PPB	12 Jul 2017 15:12:41	U	SMPL	
I4140-02 - 1	EP-05	1289	0.2963	-7471A	PPB	12 Jul 2017 15:14:58	U	SMPL	
I4140-03 - 1	EP-09	252	0.0272	-7471A	PPB	12 Jul 2017 15:17:16	U	SMPL	
I4140-04 - 1	EP-15	48	-0.0258	-7471A	PPB	12 Jul 2017 15:19:33	U	SMPL	
I4140-05 - 1	EP-22	302	0.0402	-7471A	PPB	12 Jul 2017 15:21:49	U	SMPL	
CCV02 - 1	CCV02	19330	4.9790	-7471A	PPB	12 Jul 2017 15:24:05	U	SMPL	
CCB02 - 1	CCB02	2	-0.0377	-7471A	PPB	12 Jul 2017 15:26:22	U	SMPL	
I4140-05DUP - 1	EP-22DUP	328	0.0469	-7471A	PPB	12 Jul 2017 15:28:38	U	SMPL	
I4140-05MS - 1	EP-22MS	15134	3.8899	-7471A	PPB	12 Jul 2017 15:30:54	U	SMPL	
I4140-05MSD - 1	EP-22MSD	14391	3.6970	-7471A	PPB	12 Jul 2017 15:33:10	U	SMPL	
I4144-01 - 1	TP1-0-3	10264	2.6258	-7471A	PPB	12 Jul 2017 15:35:27	U	SMPL	
I4144-02 - 1	TP2-0-3	7732	1.9687	-7471A	PPB	12 Jul 2017 15:37:44	U	SMPL	
I4144-03 - 1	TP3-0-3	8157	2.0790	-7471A	PPB	12 Jul 2017 15:40:00	U	SMPL	
I4146-01 - 1	FK-SF-03-001-A	2333	0.5673	-7471A	PPB	12 Jul 2017 15:42:18	U	SMPL	
I4146-03 - 1	FK-SF-03-001-B	2201	0.5331	-7471A	PPB	12 Jul 2017 15:44:35	U	SMPL	
I4146-05 - 1	FK-SF-03-001-C	65540	16.9730	-7471A	PPB	12 Jul 2017 15:46:53	U	SMPL	
I4147-01 - 1	F-7-11	4577	1.1498	-7471A	PPB	12 Jul 2017 15:49:09	U	SMPL	
CCV03 - 1	CCV03	19660	5.0646	-7471A	PPB	12 Jul 2017 15:51:28	U	SMPL	
CCB03 - 1	CCB03	-5	-0.0395	-7471A	PPB	12 Jul 2017 15:53:45	U	SMPL	
I4148-01 - 1	POLE-1-2-3-4-COMP	1057	0.2361	-7471A	PPB	12 Jul 2017 15:56:02	U	SMPL	
I4149-01 - 1	OR-01-071117	2204	0.5338	-7471A	PPB	12 Jul 2017 15:58:18	U	SMPL	
I4140-05LX5 - 1		74	-0.0190	-7471A	PPB	12 Jul 2017 16:00:34	U	SMPL	
I4140-05A - 1		15789	4.0599	-7471A	PPB	12 Jul 2017 16:02:51	U	SMPL	
I4146-05DLX5 - 1		12901	3.3103	-7471A	PPB	12 Jul 2017 16:05:07	U	SMPL	
CCV04 - 1	CCV04	19570	5.0413	-7471A	PPB	12 Jul 2017 16:07:24	U	SMPL	
CCB04 - 1	CCB04	-79	-0.0587	-7471A	PPB	12 Jul 2017 16:09:42	U	SMPL	