



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88670

Temp.: 23.7 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/12/2017	14:30
2	CAL2	1	Water			7.00	07/12/2017	14:31
3	CAL3	1	Water			10.03	07/12/2017	14:32
4	ICV	1	Water			7.01	07/12/2017	14:33
5	CCV1	1	Water			2.02	07/12/2017	14:35
6	I4145-11	1	Water			7.92	07/12/2017	14:36
7	I4145-11DUP	1	Water			7.93	07/12/2017	14:37
8	I4150-02	1	Water			7.27	07/12/2017	14:38
9	CCV2	1	Water			12.01	07/12/2017	14:39

2888670

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4145

WorkList ID : 100581

Date : 7/12/2017 2:13:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2017	Water	I4145-11-E	Corrosivity		PSEG01		BASEMENT-WATER	07/11/2017	9040C
07/11/2017	Water	I4150-02-F	Corrosivity	Cool 4 deg C	PSEG01		A3591	07/11/2017	9040C

Date/Time

Received by:

Relinquished by:

07-12-17 2:13 PM
Jo
CG

Date/Time

Received by:

Relinquished by:

07-12-17 4:30 PM
Jo
CG