



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88671

Temp.: 23.7 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/12/2017	14:30
2	CAL2	1	Water			7.00	07/12/2017	14:31
3	CAL3	1	Water			10.03	07/12/2017	14:32
4	ICV	1	Water			7.01	07/12/2017	14:33
5	CCV1	1	Water			2.02	07/12/2017	14:35
6	I4145-09	1	Water			7.92	07/12/2017	14:36
7	I4145-09DUP	1	Water			7.93	07/12/2017	14:37
8	I4150-01	1	Water			7.27	07/12/2017	14:38
9	CCV2	1	Water			12.01	07/12/2017	14:39

WB 88671

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4145 WorkList ID : 100582 Date : 7/12/2017 2:10:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2017	Water	I4145-09 - E	pH		PSEG01		BASEMENT-WATER	07/11/2017	9040C
07/11/2017	Water	I4150-01 - F	pH	Cool 4 deg C	PSEG01	G22	A3591	07/11/2017	9045C

Date/Time 07-12-17 2:10 PM
Received by: JP
Relinquished by: CS

Date/Time 07-12-17 11:30 AM
Received by: CS
Relinquished by: JP