


7471A

LB88680
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38712



A= 0.0000e+000

B= 2.5942e-004

C= -3.1333e-002

Rho= 0.9998344

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	-1	0.000	-1				
0.2	0.200	0.204	0.004	907	0.0 %	907				
2.5	2.500	2.490	-0.010	9718	0.0 %	9718				
5.0	5.000	5.126	0.126	19880	0.0 %	19880				
7.5	7.500	7.400	-0.100	28647	0.0 %	28647	0			
10.0	10.000	10.011	0.011	38712	0.0 %	38712				

LB88680 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 13 Jul 2017 10:15:41				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-1	-	0.00007471A	PPB		13 Jul 2017 10:26:47	S Std
0.2 - 1	50.2	907	-	0.20007471A	PPB		13 Jul 2017 10:29:05	S Std
2.5 - 1	52.5	9718	-	2.50007471A	PPB		13 Jul 2017 10:31:23	S Std
5.0 - 1	55	19880	-	5.00007471A	PPB		13 Jul 2017 10:33:40	S Std
7.5 - 1	57.5	28647	-	7.50007471A	PPB		13 Jul 2017 10:35:58	S Std
10.0 - 1	50	38712	-	10.00007471A	PPB		13 Jul 2017 10:38:15	S Std
ICV60 - 1	ICV60	15893	4.0917	-7471A	PPB		13 Jul 2017 10:40:54	U SMPL
ICB60 - 1	ICB60	-80	-0.0521	-7471A	PPB		13 Jul 2017 10:43:10	U SMPL
CCV08 - 1	CCV08	19866	5.1224	-7471A	PPB		13 Jul 2017 10:45:25	U SMPL
CCB08 - 1	CCB08	-30	-0.0391	-7471A	PPB		13 Jul 2017 10:47:41	U SMPL
CRA - 1	CRA	910	0.2047	-7471A	PPB		13 Jul 2017 10:49:58	U SMPL
HighStd - 1	HighStd	39117	10.1166	-7471A	PPB		13 Jul 2017 10:52:13	U SMPL
ChkStd - 1	ChkStd	27128	7.0063	-7471A	PPB		13 Jul 2017 10:54:29	U SMPL
PB100601BL - 1	PBS	-207	-0.0850	-7471A	PPB		13 Jul 2017 10:56:48	U SMPL
PB100601BS - 1	LCSS	14023	3.6066	-7471A	PPB		13 Jul 2017 10:59:06	U SMPL
I4145-01 - 1	TRANSFORMER-3-4-5-COMP	1455	0.3461	-7471A	PPB		13 Jul 2017 11:01:23	U SMPL
I4163-01 - 1	SU-04-071217	1319	0.3108	-7471A	PPB		13 Jul 2017 11:03:41	U SMPL
I4164-01 - 1	72-11975-COMP	748	0.1627	-7471A	PPB		13 Jul 2017 11:30:59	U SMPL
I4166-01 - 1	COMP-1-2-3	8139	2.0801	-7471A	PPB		13 Jul 2017 11:33:16	U SMPL
I4166-01DUP - 1	COMP-1-2-3DUP	8354	2.1359	-7471A	PPB		13 Jul 2017 11:35:27	U SMPL
CCV09 - 1	CCV09	19931	5.1392	-7471A	PPB		13 Jul 2017 11:37:39	U SMPL
CCB09 - 1	CCB09	165	0.0115	-7471A	PPB		13 Jul 2017 11:39:53	U SMPL
I4166-01MS - 1	COMP-1-2-3MS	21706	5.5997	-7471A	PPB		13 Jul 2017 11:42:08	U SMPL
I4166-01MSD - 1	COMP-1-2-3MSD	21517	5.5507	-7471A	PPB		13 Jul 2017 11:44:23	U SMPL
I4166-07 - 1	COMP-4-5-6	43757	11.3203	-7471A	PPB		13 Jul 2017 11:46:41	U SMPL
I4172-01 - 1	E2-P6-SSW	38666	9.9996	-7471A	PPB		13 Jul 2017 11:48:59	U SMPL
I4172-02 - 1	E2-P5-BASE	139842	36.2470	-7471A	PPB		13 Jul 2017 11:51:17	U SMPL
I4172-03 - 1	E2-P6-BASE	95541	24.7543	-7471A	PPB		13 Jul 2017 11:53:36	U SMPL
I4172-04 - 1	E2-P6-NSW	6074	1.5444	-7471A	PPB		13 Jul 2017 11:55:57	U SMPL
I4172-05 - 1	E2-P6-ESW	57708	14.9395	-7471A	PPB		13 Jul 2017 11:58:18	U SMPL
I4172-06 - 1	E2-04-BASE	49805	12.8893	-7471A	PPB		13 Jul 2017 12:00:35	U SMPL
I4172-07 - 1	E2-04-WSW	285268	73.9740	-7471A	PPB		13 Jul 2017 12:02:53	U SMPL
CCV10 - 1	CCV10	18772	4.8386	-7471A	PPB		13 Jul 2017 12:05:12	U SMPL
CCB10 - 1	CCB10	-87	-0.0539	-7471A	PPB		13 Jul 2017 12:11:14	U SMPL
I4172-08 - 1	E2-CONCRETE-BATCH-15	2366	0.5825	-7471A	PPB		13 Jul 2017 12:13:30	U SMPL
I4172-10 - 1	E2-05-BASE	14479	3.7249	-7471A	PPB		13 Jul 2017 12:15:47	U SMPL
I4172-11 - 1	E2-05-NSW	30425	7.8616	-7471A	PPB		13 Jul 2017 12:18:03	U SMPL
I4172-12 - 1	E2-05-SSW	26849	6.9339	-7471A	PPB		13 Jul 2017 12:20:19	U SMPL
I4172-13 - 1	E2-05-ESW	17539	4.5187	-7471A	PPB		13 Jul 2017 12:22:38	U SMPL
I4172-14 - 1	E2-05-WSW	27558	7.1179	-7471A	PPB		13 Jul 2017 12:24:56	U SMPL
I4166-01LX5 - 1		1407	0.3337	-7471A	PPB		13 Jul 2017 12:27:13	U SMPL
I4166-01A - 1		23517	6.0695	-7471A	PPB		13 Jul 2017 12:29:32	U SMPL
I4166-07DLX2 - 1		22239	5.7380	-7471A	PPB		13 Jul 2017 12:31:49	U SMPL
I4172-02DLX10 - 1		14149	3.6393	-7471A	PPB		13 Jul 2017 12:34:06	U SMPL
CCV11 - 1	CCV11	19918	5.1359	-7471A	PPB		13 Jul 2017 12:36:24	U SMPL
CCB11 - 1	CCB11	-127	-0.0643	-7471A	PPB		13 Jul 2017 12:38:40	U SMPL
I4172-03DLX5 - 1		19655	5.0676	-7471A	PPB		13 Jul 2017 12:40:57	U SMPL
I4172-05DLX5 - 1		11222	2.8799	-7471A	PPB		13 Jul 2017 12:43:13	U SMPL
I4172-06DLX5 - 1		9702	2.4856	-7471A	PPB		13 Jul 2017 12:45:30	U SMPL
I4172-07DLX10 - 1		32004	8.2713	-7471A	PPB		13 Jul 2017 12:47:46	U SMPL
CCV12 - 1	CCV12	20012	5.1603	-7471A	PPB		13 Jul 2017 12:52:38	U SMPL
CCB12 - 1	CCB12	-86	-0.0536	-7471A	PPB		13 Jul 2017 12:54:55	U SMPL