



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88688

Temp.: 24.2 °C

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/13/2017	12:25
2	CAL2	1	Water			7.01	07/13/2017	12:26
3	CAL3	1	Water			10.03	07/13/2017	12:27
4	ICV	1	Water			7.00	07/13/2017	12:30
5	CCV1	1	Water			2.01	07/13/2017	12:31
6	I4177-01	1	Water			4.26	07/13/2017	12:31
7	I4178-01	1	Water			4.45	07/13/2017	12:32
8	I4179-01	1	Water			4.57	07/13/2017	12:35
9	I4179-01DUP	1	Water			4.58	07/13/2017	12:36
10	CCV2	1	Water			12.02	07/13/2017	12:37

2B88688

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4177

WorkList ID : 100648

Date : 7/13/2017 12:10:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4177-01-A	pH	Cool 4 deg C	STER05	D11	POOLWATER	07/12/2017	SM 4500-PH B
	Water	I4178-01-A	pH	Cool 4 deg C	STER16	E11	MAINCELL	07/11/2017	SM 4500-PH B
	Water	I4179-01-A	pH	Cool 4 deg C	STER16	E11	EXCELL	07/11/2017	SM 4500-PH B

Date/Time 07/13/17 2:10 PM
Received by: JP
Relinquished by: JP

Date/Time 07/13/17 2:10 PM
Received by: JP
Relinquished by: JP