



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88694

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/12/2017	17:10
2	CAL2	1	Water			7.01	07/12/2017	17:11
3	CAL3	1	Water			10.04	07/12/2017	17:12
4	ICV	1	Water			7.00	07/12/2017	17:13
5	CCV1	1	Water			2.01	07/12/2017	17:14
6	I4174-01	1	Solid	20.03	20	8.25	07/12/2017	17:15
7	I4174-01DUP	1	Solid	20.04	20	8.26	07/12/2017	17:16
8	CCV2	1	Water			12.01	07/12/2017	17:17

2088694

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4174 WorkList ID : 100651 Date : 7/12/2017 16:50:14 pm

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4174-01	Corrosivity	Cool 4 deg C	GUAR01	D11	TL-WTP-02	06/22/2017	9045C

Date/Time 07-12-17 16:50 PM
Received by: [Signature]
Relinquished by: _____

Date/Time 07-12-17 5:35 PM
Received by: [Signature]
Relinquished by: [Signature]