


ISM02.4

LB 88701
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39154



A= 0.0000e+000

B= 2.5663e-004

C= 3.6676e-002

Rho= 0.9998613

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.048	0.048	46	0.000	46				
0.2	0.200	0.229	0.029	749	0.0 %	749				
2.5	2.500	2.436	-0.064	9349	0.0 %	9349				
5.0	5.000	4.992	-0.008	19310	0.0 %	19310				
7.5	7.500	7.410	-0.090	28731	0.0 %	28731				
10.0	10.000	10.085	0.085	39154	0.0 %	39154				

0.00
 15
 -3
 0
 -1
 1
 203
 7113117

LB88701
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	46	-		0 ISM02.4	PPB	7/13/2017 16:16	Std
	0.2 S0.2	749	-		0.2 ISM02.4	PPB	7/13/2017 16:18	Std
	2.5 S2.5	9349	-		2.5 ISM02.4	PPB	7/13/2017 16:20	Std
	5 S5	19310	-		5 ISM02.4	PPB	7/13/2017 16:23	Std
	7.5 S7.5	28731	-		7.5 ISM02.4	PPB	7/13/2017 16:25	Std
	10 S10	39154	-		10 ISM02.4	PPB	7/13/2017 16:30	Std
ICV	ICV	14805	3.8361	-	ISM02.4	PPB	7/13/2017 16:33	SMPL
ICB	ICB	-141	0.0005	-	ISM02.4	PPB	7/13/2017 16:35	SMPL
CCV018	CCV018	19427	5.0222	-	ISM02.4	PPB	7/13/2017 16:37	SMPL
CCB018	CCB018	-116	0.0069	-	ISM02.4	PPB	7/13/2017 16:39	SMPL
PB100600BL	PBW003	-19	0.0318	-	ISM02.4	PPB	7/13/2017 16:42	SMPL
I4110-01	MBDWW6	63146	16.2417	-	ISM02.4	PPB	7/13/2017 16:44	SMPL
I4110-02	MBDWW7	547224	HIGH	-	ISM02.4	PPB	7/13/2017 16:46	
I4110-03	MBDWW8	675497	HIGH	-	ISM02.4	PPB	7/13/2017 16:48	
I4110-04	MBDWW9	512780	HIGH	-	ISM02.4	PPB	7/13/2017 16:50	
I4110-05	MBDWX0	20959	5.4153	-	ISM02.4	PPB	7/13/2017 16:53	SMPL
I4110-06	MBDWX1	253970	65.2126	-	ISM02.4	PPB	7/13/2017 16:56	SMPL
I4110-07	MBDWX2	9494	2.4731	-	ISM02.4	PPB	7/13/2017 16:58	SMPL
I4110-08	MBDWX3	320645	HIGH	-	ISM02.4	PPB	7/13/2017 17:01	
I4110-09	MBDWX4	34638	8.9258	-	ISM02.4	PPB	7/13/2017 17:02	SMPL
I4110-10	MBDWX5	10217	2.6586	-	ISM02.4	PPB	7/13/2017 17:05	SMPL
I4110-11	MBDWX6	211146	54.2227	-	ISM02.4	PPB	7/13/2017 17:07	SMPL
I4110-12	MBDWX7	4826	1.2752	-	ISM02.4	PPB	7/13/2017 17:09	SMPL
CCV019	CCV019	19028	4.9198	-	ISM02.4	PPB	7/13/2017 17:12	SMPL
CCB019	CCB019	-236	-0.0239	-	ISM02.4	PPB	7/13/2017 17:14	SMPL
I4110-13	MBDWX8	14076	3.649	-	ISM02.4	PPB	7/13/2017 17:17	SMPL
I4110-14	MBDWX9	99101	25.4688	-	ISM02.4	PPB	7/13/2017 17:19	SMPL
I4110-15	MBDWY0	466606	HIGH	-	ISM02.4	PPB	7/13/2017 17:21	
I4110-16	MBDWY0D	472132	HIGH	-	ISM02.4	PPB	7/13/2017 17:23	
I4110-17	MBDWY0S	470521	HIGH	-	ISM02.4	PPB	7/13/2017 17:25	
PB100621BL	PBW003	-240	-0.0249	-	ISM02.4	PPB	7/13/2017 17:30	SMPL
I3943-01	MBDC79	150	0.0752	-	ISM02.4	PPB	7/13/2017 17:32	SMPL
I3943-02	MBDC79D	-65	0.02	-	ISM02.4	PPB	7/13/2017 17:37	SMPL
I3943-03	MBDC79S	3839	1.0219	-	ISM02.4	PPB	7/13/2017 17:40	SMPL
CCV020	CCV020	19117	4.9426	-	ISM02.4	PPB	7/13/2017 17:42	SMPL
CCB020	CCB020	-161	-0.0046	-	ISM02.4	PPB	7/13/2017 17:44	SMPL
I4110-01DLX5	MBDWW6	12216	3.1716	-	ISM02.4	PPB	7/13/2017 17:46	SMPL
I4110-02DLX10	MBDWW7	85085	21.8719	-	ISM02.4	PPB	7/13/2017 17:49	SMPL
I4110-03DLX10	MBDWW8	124099	31.884	-	ISM02.4	PPB	7/13/2017 17:51	SMPL
I4110-04DLX10	MBDWW9	79579	20.4589	-	ISM02.4	PPB	7/13/2017 17:53	SMPL
I4110-06DLX10	MBDWX1	26995	6.9644	-	ISM02.4	PPB	7/13/2017 17:56	SMPL
I4110-08DLX10	MBDWX3	42311	10.8949	-	ISM02.4	PPB	7/13/2017 17:58	SMPL
I4110-14DLX5	MBDWX9	18889	4.8841	-	ISM02.4	PPB	7/13/2017 18:03	SMPL
I4110-15DLX10	MBDWY0	69290	17.8185	-	ISM02.4	PPB	7/13/2017 18:05	SMPL
I4110-16DLX10	MBDWY0D	68769	17.6847	-	ISM02.4	PPB	7/13/2017 18:07	SMPL
I4110-17DLX10	MBDWY0S	70265	18.0687	-	ISM02.4	PPB	7/13/2017 18:09	SMPL

LB88701

INSTRUMENT ID : CV1

I4110-02DL2X50	MBDWW7	17870	4.6226 -	ISM02.4	PPB	7/13/2017 18:12	SMPL
I4110-03DL2X50	MBDWW8	25482	6.5761 -	ISM02.4	PPB	7/13/2017 18:14	SMPL
I4110-04DL2X50	MBDWW9	16137	4.1779 -	ISM02.4	PPB	7/13/2017 18:16	SMPL
I4110-08DL2X50	MBDWX3	8554	2.2319 -	ISM02.4	PPB	7/13/2017 18:19	SMPL
I4110-15DL2X50	MBDWY0	14040	3.6397 -	ISM02.4	PPB	7/13/2017 18:21	SMPL
I4110-16DL2X50	MBDWY0D	13834	3.5869 -	ISM02.4	PPB	7/13/2017 18:23	SMPL
I4110-17DL2X50	MBDWY0S	13934	3.6125 -	ISM02.4	PPB	7/13/2017 18:26	SMPL
I4110-11DLX10	MBDWX6	22159	5.7233 -	ISM02.4	PPB	7/13/2017 18:28	SMPL
CCV021	CCV021	19247	4.976 -	ISM02.4	PPB	7/13/2017 18:30	SMPL
CCB021	CCB021	-151	-0.0021 -	ISM02.4	PPB	7/13/2017 18:32	SMPL