



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB88702

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	07/13/2017	13:20
2	I4145-11		1	>100.0	>212.0	07/13/2017	13:50
3	I4145-11DUP		1	>100.0	>212.0	07/13/2017	14:15

Result = (Celsius * 1.8) + 32

LB 88702

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71317

WorkList ID : 100647

Date : 7/13/2017 12:42:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/14/2017	Water	I4145-11 C	Flash Point		PSEG01		BASEMENT-WATER	07/11/2017	1010A

Date/Time 07/13/17 12:45 PM
Received by: KS
Relinquished by: CP

Date/Time 07/13/17 14:25
Received by: CP
Relinquished by: KS