



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88706

Temp.: 24.6 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/13/2017	16:30
2	CAL2	1	Water			7.00	07/13/2017	16:31
3	CAL3	1	Water			10.03	07/13/2017	16:33
4	ICV	1	Water			7.01	07/13/2017	16:35
5	CCV1	1	Water			2.02	07/13/2017	16:36
6	I4184-01	1	Solid	20.03	20	7.86	07/13/2017	16:37
7	I4184-01DUP	1	Solid	20.04	20	7.87	07/13/2017	16:38
8	I4184-02	1	Solid	20.05	20	7.42	07/13/2017	16:40
9	I4184-03	1	Solid	20.02	20	7.59	07/13/2017	16:41
10	I4184-04	1	Solid	20.03	20	7.74	07/13/2017	16:42
11	I4184-05	1	Solid	20.03	20	8.03	07/13/2017	16:43
12	I4184-06	1	Solid	20.02	20	8.27	07/13/2017	16:45
13	I4184-07	1	Solid	20.04	20	7.31	07/13/2017	16:46
14	I4184-08	1	Solid	20.02	20	7.42	07/13/2017	16:49
15	I4184-09	1	Solid	20.01	20	8.48	07/13/2017	16:50
16	CCV2	1	Water			12.02	07/13/2017	16:51
17	I4184-10	1	Solid	20.02	20	7.54	07/13/2017	16:52
18	CCV3	1	Water			12.02	07/13/2017	17:00

WORKLIST(Hardcopy Internal Chain)

LB 88706

WorkList Name : PH I4184

WorkList ID : 100657

Date : 7/13/2017 4:20:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2017	Solid	I4184-01-C	pH	Cool 4 deg C	WEST04	G41	MC0B01	07/11/2017	9045C
07/11/2017	Solid	I4184-02-B	pH	Cool 4 deg C	WEST04	G41	MC0B02	07/11/2017	9045C
07/11/2017	Solid	I4184-03-C	pH	Cool 4 deg C	WEST04	G41	MC0B03	07/11/2017	9045C
07/11/2017	Solid	I4184-04-B	pH	Cool 4 deg C	WEST04	G41	MC0B04	07/11/2017	9045C
07/12/2017	Solid	I4184-05-C	pH	Cool 4 deg C	WEST04	G41	MC0B23	07/12/2017	9045C
07/12/2017	Solid	I4184-06-C	pH	Cool 4 deg C	WEST04	G41	MC0B78	07/12/2017	9045C
07/12/2017	Solid	I4184-07-C	pH	Cool 4 deg C	WEST04	G41	MC0B79	07/12/2017	9045C
07/12/2017	Solid	I4184-08-B	pH	Cool 4 deg C	WEST04	G41	MC0B80	07/12/2017	9045C
07/12/2017	Solid	I4184-09-C	pH	Cool 4 deg C	WEST04	G41	MC0B82	07/12/2017	9045C
07/12/2017	Solid	I4184-10-B	pH	Cool 4 deg C	WEST04	G41	MC0B83	07/12/2017	9045C

Date/Time 07-13-17 4:20 PM
 Received by: JA
 Relinquished by: JC

Date/Time 07-13-17 6:30 PM
 Received by: W
 Relinquished by: J4