



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88708

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/13/2017	18:00
2	CAL2	1	Water			7.01	07/13/2017	18:01
3	CAL3	1	Water			10.03	07/13/2017	18:02
4	ICV	1	Water			7.00	07/13/2017	18:03
5	CCV1	1	Water			2.01	07/13/2017	18:05
6	I4196-01	1	Solid	20.03	20	6.77	07/13/2017	18:06
7	I4196-01DUP	1	Solid	20.04	20	6.78	07/13/2017	18:08
8	CCV2	1	Water			12.02	07/13/2017	18:10

2988708

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4196

WorkList ID : 100660

Date : 7/13/2017 5:44:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4196-01 - 17	pH	Cool 4 deg C	TULL02	F11	FK-PS-01-012	07/13/2017	9045C

Date/Time 07-13-17 5:44 PM
 Received by: JD
 Relinquished by: CD

Date/Time 07-13-17 7:00 PM
 Received by: CD
 Relinquished by: JD