



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88721

Temp.: 23.8 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/14/2017	14:30
2	CAL2	1	Water			7.01	07/14/2017	14:31
3	CAL3	1	Water			10.03	07/14/2017	14:32
4	ICV	1	Water			6.99	07/14/2017	14:35
5	CCV1	1	Water			2.01	07/14/2017	14:36
6	I4203-01	1	Solid	20.03	20	7.02	07/14/2017	14:38
7	I4203-02	1	Solid	20.04	20	6.44	07/14/2017	14:39
8	I4203-03	1	Solid	20.02	20	6.38	07/14/2017	14:40
9	I4203-04	1	Solid	20.03	20	8.50	07/14/2017	14:41
10	I4203-05	1	Solid	20.02	20	8.16	07/14/2017	14:42
11	I4203-05DUP	1	Solid	20.03	20	8.17	07/14/2017	14:43
12	I4203-08	1	Solid	20.05	20	6.81	07/14/2017	14:45
13	I4203-09	1	Solid	20.04	20	7.94	07/14/2017	14:46
14	I4203-10	1	Solid	20.02	20	7.00	07/14/2017	14:47
15	I4203-11	1	Solid	20.01	20	8.20	07/14/2017	14:48
16	CCV2	1	Water			12.01	07/14/2017	14:50
17	I4203-12	1	Solid	20.02	20	7.52	07/14/2017	14:51
18	CCV3	1	Water			12.02	07/14/2017	14:53

1388f21

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4203

WorkList ID : 100696

Date : 7/14/2017 2:11:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4203-01 -F	pH	Cool 4 deg C	BEMS01	F41	DL-SB02A	07/13/2017	9045C
	Solid	I4203-02 -F	pH	Cool 4 deg C	BEMS01	F41	DL-SB02C	07/13/2017	9045C
	Solid	I4203-03 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB02A	07/13/2017	9045C
	Solid	I4203-04 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB02B	07/13/2017	9045C
	Solid	I4203-05 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB04A	07/13/2017	9045C
	Solid	I4203-08 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB04B	07/13/2017	9045C
	Solid	I4203-09 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB01A	07/13/2017	9045C
	Solid	I4203-10 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB01B	07/13/2017	9045C
	Solid	I4203-11 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB03A	07/13/2017	9045C
	Solid	I4203-12 -F	pH	Cool 4 deg C	BEMS01	F41	NB-SB03B	07/13/2017	9045C

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: