



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88722

Temp.: 21.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/14/2017	14:30
2	CAL2	1	Water			7.01	07/14/2017	14:31
3	CAL3	1	Water			10.03	07/14/2017	14:32
4	ICV	1	Water			6.99	07/14/2017	14:33
5	CCV1	1	Water			2.02	07/14/2017	15:10
6	I4203-13	1	Water			5.99	07/14/2017	15:11
7	I4203-13DUP	1	Water			6.00	07/14/2017	15:13
8	CCV2	1	Water			12.01	07/14/2017	15:15

LB 88722

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHW14203 WorkList ID : 100697 Date : 7/14/2017 2:12:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4203-13	pH	Cool 4 deg C	BEMS01	F41	EB-2017-07-13	07/13/2017	SM 4500-PH B

Date/Time 6/14/17 2:12 PM
Received by: JR
Relinquished by: JR

Date/Time 07-14-17 5:20 PM
Received by: JR
Relinquished by: JR