



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88728

Temp.: 23.8 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/14/2017	15:30
2	CAL2	1	Water			6.99	07/14/2017	15:31
3	CAL3	1	Water			10.02	07/14/2017	15:32
4	ICV	1	Water			7.01	07/14/2017	15:33
5	CCV1	1	Water			2.02	07/14/2017	15:35
6	I4203-14	1	Solid	20.03	20	7.59	07/14/2017	15:36
7	I4203-14DUP	1	Solid	20.02	20	7.60	07/14/2017	15:37
8	I4203-15	1	Solid	20.01	20	8.59	07/14/2017	15:39
9	I4203-16	1	Solid	20.04	20	8.39	07/14/2017	15:40
10	CCV2	1	Water			12.01	07/14/2017	15:42

LB 88728

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I4203

WorkList ID : 100693

Date : 7/14/2017 1:05:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4203-14	Corrosivity	Cool 4 deg C	BEMS01	F41	DL-SB02A	07/13/2017	9045C
	Solid	I4203-15	Corrosivity	Cool 4 deg C	BEMS01	F41	NB-SB02B	07/13/2017	9045C
	Solid	I4203-16	Corrosivity	Cool 4 deg C	BEMS01	F41	NB-SB03A	07/13/2017	9045C

Date/Time 07-14-17 1:05 PM
 Received by: JJ
 Relinquished by: JR

Date/Time 07-14-17 5:00 PM
 Received by: JR
 Relinquished by: JJ