


7470A

LB 88742
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33123



A= 0.0000e+000

B= 3.0326e-004

C= -1.0158e-002

Rho= 0.9999294

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	62	0.000	62				
0.2	0.200	0.186	-0.014	647	0.0 %	647				
2.5	2.500	2.500	-0.000	8276	0.0 %	8276				
5.0	5.000	5.055	0.055	16702	0.0 %	16702				
7.5	7.500	7.416	-0.084	24488	0.0 %	24488				
10.0	10.000	10.035	0.035	33123	0.0 %	33123				

LB88742 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 17 Jul 2017 15:20:24				
Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	62	-	0.0000	7470A	PPB	17 Jul 2017 15:25:19	S Std
0.2 - 1	50.2	647	-	0.2000	7470A	PPB	17 Jul 2017 15:27:28	S Std
2.5 - 1	52.5	8276	-	2.5000	7470A	PPB	17 Jul 2017 15:29:36	S Std
5.0 - 1	55	16702	-	5.0000	7470A	PPB	17 Jul 2017 15:31:44	S Std
7.5 - 1	57.5	24488	-	7.5000	7470A	PPB	17 Jul 2017 15:33:52	S Std
10.0 - 1	510	33123	-	10.0000	7470A	PPB	17 Jul 2017 15:36:00	S Std
ICV62 - 1	ICV62	12793	3.8694	-	7470A	PPB	17 Jul 2017 15:38:35	U SMPL
ICB62 - 1	ICB62	-151	-0.0560	-	7470A	PPB	17 Jul 2017 15:40:41	U SMPL
CCV22 - 1	CCV22	16792	5.0822	-	7470A	PPB	17 Jul 2017 15:42:48	U SMPL
CCB22 - 1	CCB22	-169	-0.0614	-	7470A	PPB	17 Jul 2017 15:44:55	U SMPL
CRA - 1	CRA	652	0.1876	-	7470A	PPB	17 Jul 2017 15:47:02	U SMPL
HighStd - 1	HighStd	33639	10.1912	-	7470A	PPB	17 Jul 2017 15:49:09	U SMPL
ChkStd - 1	ChkStd	23654	7.1632	-	7470A	PPB	17 Jul 2017 15:51:16	U SMPL
PB100679BL - 1	PBW	-218	-0.0763	-	7470A	PPB	17 Jul 2017 15:53:24	U SMPL
PB100679BS - 1	LCSW	12586	3.8067	-	7470A	PPB	17 Jul 2017 15:55:33	U SMPL
PB100679TB - 1	PB100627TB	-140	-0.0526	-	7470A	PPB	17 Jul 2017 15:57:41	U SMPL
I4176-05 - 1	VNJ-225-241-72-11938-COMP	130	0.0293	-	7470A	PPB	17 Jul 2017 15:59:49	U SMPL
I4176-08 - 1	72-12003	-59	-0.0281	-	7470A	PPB	17 Jul 2017 16:01:57	U SMPL
I4185-02 - 1	HDD-PILWC1	-49	-0.0250	-	7470A	PPB	17 Jul 2017 16:04:06	U SMPL
I4188-03 - 1	B1(0-10)COMP	-35	-0.0208	-	7470A	PPB	17 Jul 2017 16:06:12	U SMPL
CCV23 - 1	CCV23	16391	4.9606	-	7470A	PPB	17 Jul 2017 16:08:19	U SMPL
CCB23 - 1	CCB23	-132	-0.0502	-	7470A	PPB	17 Jul 2017 16:10:26	U SMPL
I4193-03 - 1	TR-04-071317	-42	-0.0229	-	7470A	PPB	17 Jul 2017 16:12:32	U SMPL
I4201-04 - 1	TP-2-RIVER-ROAD	-49	-0.0250	-	7470A	PPB	17 Jul 2017 16:14:39	U SMPL
I4201-04DUP - 1	TP-2-RIVER-ROADDUP	-18	-0.0156	-	7470A	PPB	17 Jul 2017 16:16:46	U SMPL
I4201-04MS - 1	TP-2-RIVER-ROADMS	12041	3.6414	-	7470A	PPB	17 Jul 2017 16:18:53	U SMPL
I4201-04MSD - 1	TP-2-RIVER-ROADMSD	12034	3.6393	-	7470A	PPB	17 Jul 2017 16:21:00	U SMPL
PB100680BL - 1	PBW	-131	-0.0499	-	7470A	PPB	17 Jul 2017 16:23:08	U SMPL
PB100680BS - 1	LCSW	12129	3.6681	-	7470A	PPB	17 Jul 2017 16:25:16	U SMPL
PB100680TB - 1	PB100658TB	-105	-0.0420	-	7470A	PPB	17 Jul 2017 16:27:24	U SMPL
I4203-14 - 1	DL-SB02A	-42	-0.0229	-	7470A	PPB	17 Jul 2017 16:29:33	U SMPL
I4203-14DUP - 1	DL-SB02ADUP	-34	-0.0205	-	7470A	PPB	17 Jul 2017 16:31:40	U SMPL
CCV24 - 1	CCV24	16651	5.0394	-	7470A	PPB	17 Jul 2017 16:33:47	U SMPL
CCB24 - 1	CCB24	-141	-0.0529	-	7470A	PPB	17 Jul 2017 16:35:53	U SMPL
I4203-14MS - 1	DL-SB02AMS	12015	3.6335	-	7470A	PPB	17 Jul 2017 16:38:01	U SMPL
I4203-14MSD - 1	DL-SB02AMSD	12156	3.6763	-	7470A	PPB	17 Jul 2017 16:40:08	U SMPL
I4203-15 - 1	NB-SB02B	-123	-0.0475	-	7470A	PPB	17 Jul 2017 16:42:15	U SMPL
I4203-16 - 1	NB-SB03A	-19	-0.0159	-	7470A	PPB	17 Jul 2017 16:44:22	U SMPL
I4211-01 - 1	I16-005-MH	-3	-0.0111	-	7470A	PPB	17 Jul 2017 16:46:29	U SMPL
PB100681BL - 1	PBW	-25	-0.0177	-	7470A	PPB	17 Jul 2017 16:48:37	U SMPL
PB100681BS - 1	LCSW	12976	3.9249	-	7470A	PPB	17 Jul 2017 16:50:45	U SMPL
I4203-13 - 1	EB-2017-07-13	-127	-0.0487	-	7470A	PPB	17 Jul 2017 16:52:53	U SMPL
I4205-01 - 1	NB-TWP01	367	0.1011	-	7470A	PPB	17 Jul 2017 16:55:02	U SMPL
I4205-01DUP - 1	NB-TWP01DUP	196	0.0493	-	7470A	PPB	17 Jul 2017 16:57:09	U SMPL
CCV25 - 1	CCV25	16420	4.9694	-	7470A	PPB	17 Jul 2017 17:02:02	U SMPL
CCB25 - 1	CCB25	-146	-0.0544	-	7470A	PPB	17 Jul 2017 17:04:09	U SMPL
I4205-01MS - 1	NB-TWP01MS	13211	3.9962	-	7470A	PPB	17 Jul 2017 17:06:17	U SMPL
I4205-01MSD - 1	NB-TWP01MSD	13183	3.9877	-	7470A	PPB	17 Jul 2017 17:08:24	U SMPL
I4205-02 - 1	NB-TWP07	-34	-0.0205	-	7470A	PPB	17 Jul 2017 17:10:31	U SMPL
PB100682BL - 1	PBW	-43	-0.0232	-	7470A	PPB	17 Jul 2017 17:15:08	U SMPL
PB100682BS - 1	LCSW	13429	4.0623	-	7470A	PPB	17 Jul 2017 17:17:15	U SMPL
I4184-01 - 1	MC0B01	243	0.0635	-	7470A	PPB	17 Jul 2017 17:19:23	U SMPL
I4184-01DUP - 1	MC0B01DUP	99	0.0199	-	7470A	PPB	17 Jul 2017 17:21:31	U SMPL
I4184-01MS - 1	MC0B01MS	15191	4.5967	-	7470A	PPB	17 Jul 2017 17:23:40	U SMPL
I4184-01MSD - 1	MC0B01MSD	15214	4.6036	-	7470A	PPB	17 Jul 2017 17:25:48	U SMPL
I4184-02 - 1	MC0B02	12	-0.0065	-	7470A	PPB	17 Jul 2017 17:27:55	U SMPL
CCV26 - 1	CCV26	16905	5.1165	-	7470A	PPB	17 Jul 2017 17:30:03	U SMPL
CCB26 - 1	CCB26	-161	-0.0590	-	7470A	PPB	17 Jul 2017 17:32:10	U SMPL
I4184-03 - 1	MC0B03	179	0.0441	-	7470A	PPB	17 Jul 2017 17:34:18	U SMPL
I4184-04 - 1	MC0B04	1611	0.4784	-	7470A	PPB	17 Jul 2017 17:36:25	U SMPL
I4184-05 - 1	MC0B23	0	-0.0102	-	7470A	PPB	17 Jul 2017 17:38:32	U SMPL
I4184-06 - 1	MC0B78	9	-0.0074	-	7470A	PPB	17 Jul 2017 17:40:39	U SMPL
I4184-07 - 1	MC0B79	155	0.0368	-	7470A	PPB	17 Jul 2017 17:42:46	U SMPL
I4184-08 - 1	MC0B80	20	-0.0041	-	7470A	PPB	17 Jul 2017 17:44:54	U SMPL
I4184-09 - 1	MC0B82	260	0.0687	-	7470A	PPB	17 Jul 2017 17:47:02	U SMPL
I4184-10 - 1	MC0B83	50	0.0050	-	7470A	PPB	17 Jul 2017 17:49:10	U SMPL
CCV27 - 1	CCV27	16575	5.0164	-	7470A	PPB	17 Jul 2017 17:54:36	U SMPL
CCB27 - 1	CCB27	-166	-0.0605	-	7470A	PPB	17 Jul 2017 17:56:43	U SMPL
I4201-04LX5 - 1		-41	-0.0226	-	7470A	PPB	17 Jul 2017 17:58:51	U SMPL
I4201-04A - 1		13284	4.0183	-	7470A	PPB	17 Jul 2017 18:00:58	U SMPL
I4203-14LX5 - 1		-151	-0.0560	-	7470A	PPB	17 Jul 2017 18:03:06	U SMPL
I4203-14A - 1		12939	3.9137	-	7470A	PPB	17 Jul 2017 18:05:13	U SMPL
I4205-01LX5 - 1		-76	-0.0332	-	7470A	PPB	17 Jul 2017 18:07:21	U SMPL
I4205-01A - 1		14013	4.2394	-	7470A	PPB	17 Jul 2017 18:09:28	U SMPL
I4184-01LX5 - 1		-103	-0.0414	-	7470A	PPB	17 Jul 2017 18:11:36	U SMPL
I4184-01A - 1		15214	4.6036	-	7470A	PPB	17 Jul 2017 18:13:43	U SMPL
CCV28 - 1	CCV28	16598	5.0234	-	7470A	PPB	17 Jul 2017 18:15:52	U SMPL
CCB28 - 1	CCB28	-154	-0.0569	-	7470A	PPB	17 Jul 2017 18:18:00	U SMPL