

Analysis Method: 9034

ANALYST: ketankumar

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB88747

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2206
IODINE SOLUTION .025N 1L	W2251
Starch Solution, 4L	W2121

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB100686BL		1	5.02	50	2.00	0.00	1.70	1.70	0.30	0.00	0.00	07/17/2017	13:30
2	I4203-14		1	5.04	50	2.00	1.70	3.38	1.68	0.32	0.02	1.59	07/17/2017	13:33
3	I4203-14DUP		1	5.03	50	2.00	3.38	5.08	1.70	0.30	0.00	0.00	07/17/2017	13:35
4	I4203-15		1	5.04	50	2.00	5.08	6.70	1.62	0.38	0.08	6.35	07/17/2017	13:38
5	I4203-16		1	5.05	50	2.00	6.70	8.16	1.46	0.54	0.24	19.01	07/17/2017	13:40

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume