

CHEMTECH

284 Sheffield Street,

Mountainside, NJ 07092

Reviewed by : AB

7/17/2017 11:22


Test Total CN

Accepted 7/17/2017 11:22

Factor 1007
Bias 0.003

Coeff. of det. 0.999960

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	1.6138	0.0000	
2	5.0PPBCN	0.008	5.2891	5.0000	
3	10PPBCN	0.013	9.9080	10.0000	
4	50PPBCN	0.051	48.1871	50.0000	
5	100PPBCN	0.102	99.0475	100.0000	
6	250PPBCN	0.253	251.1677	250.0000	
7	500PPBCN	0.500	499.7869	500.0000	

=====

Test results

Aquakem 7.2AQ1

UB86849
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

7/17/2017 16:06

Test: ^{Reactive} Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.218	0.0	0.097	
ICB1	1.215	0.0	0.004	
CCV1	242.835	0.0	0.244	
CCB1	1.614	0.0	0.005	
PB100685BL	1.441	0.0	0.005	
I4203-14	0.478	0.0	0.004	
I4203-14DUP	0.589	0.0	0.004	
I4203-15	0.404	0.0	0.004	
I4203-16	0.770	0.0	0.004	
CCV2	244.602	0.0	0.246	
CCB2	1.688	0.0	0.005	

N	11
Mean	53.623
SD	98.0083
CV%	182.77

LB88949

H

8/12/18

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 17 16:04:58 2017

Mon Jul 17 16:05:08 2017

Sample Id	Sam/Ctr/c	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	1.6138	µg/l	7/17/2017 11:21:34	
5.0PPBCN	A	Reactive CN	P	5.2891	µg/l	7/17/2017 11:21:35	
10PPBCN	A	Reactive CN	P	9.908	µg/l	7/17/2017 11:21:36	
50PPBCN	A	Reactive CN	P	48.1871	µg/l	7/17/2017 11:21:37	
100PPBCN	A	Reactive CN	P	99.0475	µg/l	7/17/2017 11:21:38	
250PPBCN	A	Reactive CN	P	251.1677	µg/l	7/17/2017 11:21:39	
500PPBCN	A	Reactive CN	P	499.7869	µg/l	7/17/2017 11:21:40	
ICV1	S	Reactive CN	P	94.2182	µg/l	7/17/2017 16:04:58	
ICB1	S	Reactive CN	P	1.2155	µg/l	7/17/2017 16:04:59	
CCV1	S	Reactive CN	P	242.8345	µg/l	7/17/2017 16:05:00	
CCB1	S	Reactive CN	P	1.6142	µg/l	7/17/2017 16:05:01	
PB100685BL	S	Reactive CN	P	1.4409	µg/l	7/17/2017 16:05:02	
I4203-14	S	Reactive CN	P	0.4779	µg/l	7/17/2017 16:05:03	
I4203-14DUFS	S	Reactive CN	P	0.5891	µg/l	7/17/2017 16:05:04	
I4203-15	S	Reactive CN	P	0.4036	µg/l	7/17/2017 16:05:05	
I4203-16	S	Reactive CN	P	0.7698	µg/l	7/17/2017 16:05:06	
CCV2	S	Reactive CN	P	244.6019	µg/l	7/17/2017 16:05:07	
CCB2	S	Reactive CN	P	1.688	µg/l	7/17/2017 16:05:08	