



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88754

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/17/2017	18:10
2	CAL2	1	Water			7.00	07/17/2017	18:11
3	CAL3	1	Water			10.03	07/17/2017	18:12
4	ICV	1	Water			7.01	07/17/2017	18:14
5	CCV1	1	Water			2.01	07/17/2017	18:15
6	I4216-07	1	Solid	20.03	20	8.10	07/17/2017	18:16
7	I4216-07DUP	1	Solid	20.04	20	8.11	07/17/2017	18:17
8	CCV2	1	Water			12.01	07/17/2017	18:20

2B88754

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4216

WorkList ID : 100754

Date : 7/17/2017 6:03:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4216-07	pH	Cool 4 deg C	PARS02	F31	VITALE-TOPSOIL-1A	07/13/2017	9045C

Date/Time 07-17-17
 Received by: JO
 Relinquished by: _____

Date/Time 07-17-17
 Received by: _____
 Relinquished by: JO