


7471A

LB 88758
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38085



A= 0.0000e+000

B= 2.6202e-004

C= 2.7711e-002

Rho= 0.9999615

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	26	0.000	26				
0.2	0.200	0.202	0.002	665	0.0 %	665				
2.5	2.500	2.479	-0.021	9355	0.0 %	9355				
5.0	5.000	4.942	-0.058	18756	0.0 %	18756				
7.5	7.500	7.535	0.035	28653	0.0 %	28653	0			
10.0	10.000	10.007	0.007	38085	0.0 %	38085				

LB88758 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 18 Jul 2017 09:05:54

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	5.0	26	-	0.0007471A	PPB	18 Jul 2017 09:15:16	S	Std
0.2 - 1	50.2	665	-	0.20007471A	PPB	18 Jul 2017 09:17:33	S	Std
2.5 - 1	52.5	9355	-	2.50007471A	PPB	18 Jul 2017 09:19:51	S	Std
5.0 - 1	55	18756	-	5.00007471A	PPB	18 Jul 2017 09:22:08	S	Std
7.5 - 1	57.5	28653	-	7.50007471A	PPB	18 Jul 2017 09:24:26	S	Std
10.0 - 1	510	38085	-	10.00007471A	PPB	18 Jul 2017 09:29:17	S	Std
ICV63 - 1	ICV63	14724	3.8857	-7471A	PPB	18 Jul 2017 09:32:00	U	SMPL
ICB63 - 1	ICB63	-267	-0.0422	-7471A	PPB	18 Jul 2017 09:34:16	U	SMPL
CCV29 - 1	CCV29	20219	5.3256	-7471A	PPB	18 Jul 2017 09:36:33	U	SMPL
CCB29 - 1	CCB29	-153	-0.0124	-7471A	PPB	18 Jul 2017 09:38:49	U	SMPL
CRA - 1	CRA	738	0.2211	-7471A	PPB	18 Jul 2017 09:41:06	U	SMPL
HighStd - 1	HighStd	38938	10.2304	-7471A	PPB	18 Jul 2017 09:43:22	U	SMPL
ChkStd - 1	ChkStd	27371	7.1996	-7471A	PPB	18 Jul 2017 09:45:38	U	SMPL
PB100698BL - 1	PBS	-242	-0.0357	-7471A	PPB	18 Jul 2017 09:47:56	U	SMPL
PB100698BS - 1	LCSS	13924	3.6761	-7471A	PPB	18 Jul 2017 09:50:15	U	SMPL
I4176-01 - 1	VNJ-225-241-72-11938-COMP	8775	2.3270	-7471A	PPB	18 Jul 2017 09:52:32	U	SMPL
I4176-06 - 1	72-12003	1996	0.5507	-7471A	PPB	18 Jul 2017 09:54:50	U	SMPL
I4185-01 - 1	HDD-PILWC1	62	0.0440	-7471A	PPB	18 Jul 2017 09:57:08	U	SMPL
I4187-01 - 1	CI-SB1-0-2	3438	0.9285	-7471A	PPB	18 Jul 2017 09:59:25	U	SMPL
I4187-02 - 1	CI-SB2-0-2	29946	7.8743	-7471A	PPB	18 Jul 2017 10:01:41	U	SMPL
CCV30 - 1	CCV30	19465	5.1280	-7471A	PPB	18 Jul 2017 10:03:58	U	SMPL
CCB30 - 1	CCB30	-143	-0.0098	-7471A	PPB	18 Jul 2017 10:06:14	U	SMPL
I4187-03 - 1	CI-SB3-0-2	28699	7.5475	-7471A	PPB	18 Jul 2017 10:08:32	U	SMPL
I4187-04 - 1	CI-SB4-0-2	53579	14.0667	-7471A	PPB	18 Jul 2017 10:10:48	U	SMPL
I4187-05 - 1	CI-SB5-0-2	1275	0.3618	-7471A	PPB	18 Jul 2017 10:13:06	U	SMPL
I4187-06 - 1	CI-SB6-0-2	3060	0.8295	-7471A	PPB	18 Jul 2017 10:15:25	U	SMPL
I4187-07 - 1	CI-SB7-0-2	11625	3.0737	-7471A	PPB	18 Jul 2017 10:17:42	U	SMPL
I4187-08 - 1	CI-SB8-0-2	15512	4.0922	-7471A	PPB	18 Jul 2017 10:19:58	U	SMPL
I4187-09 - 1	CI-SB9-0-2	7196	1.9132	-7471A	PPB	18 Jul 2017 10:22:16	U	SMPL
I4187-09DUP - 1	CI-SB9-0-2DUP	7257	1.9292	-7471A	PPB	18 Jul 2017 10:24:34	U	SMPL
I4187-09MS - 1	CI-SB9-0-2MS	23122	6.0862	-7471A	PPB	18 Jul 2017 10:26:52	U	SMPL
I4187-09MSD - 1	CI-SB9-0-2MSD	21175	5.5761	-7471A	PPB	18 Jul 2017 10:29:08	U	SMPL
CCV31 - 1	CCV31	19662	5.1796	-7471A	PPB	18 Jul 2017 10:31:25	U	SMPL
CCB31 - 1	CCB31	-40	0.0172	-7471A	PPB	18 Jul 2017 10:33:43	U	SMPL
I4188-02 - 1	BI(0-10)COMP	5370	1.4348	-7471A	PPB	18 Jul 2017 10:35:59	U	SMPL
I4193-01 - 1	TR-04-071317	1584	0.4428	-7471A	PPB	18 Jul 2017 10:38:16	U	SMPL
I4196-01 - 1	FK-PS-01-012	392	0.1304	-7471A	PPB	18 Jul 2017 10:40:32	U	SMPL
I4201-01 - 1	TP-2-RIVER-ROAD	4906	1.3132	-7471A	PPB	18 Jul 2017 10:42:49	U	SMPL
I4216-01 - 1	VITALE-FILL-1A	1096	0.3149	-7471A	PPB	18 Jul 2017 10:45:05	U	SMPL
I4216-07 - 1	VITALE-TOPSOIL-1A	2243	0.6154	-7471A	PPB	18 Jul 2017 10:47:23	U	SMPL
I4217-01 - 1	CLEANSAND-2	169	0.0720	-7471A	PPB	18 Jul 2017 10:49:40	U	SMPL
PB100699BL - 1	PBS	-66	0.0104	-7471A	PPB	18 Jul 2017 10:51:58	U	SMPL
PB100699BS - 1	LCSS	14002	3.6966	-7471A	PPB	18 Jul 2017 10:54:16	U	SMPL
I4203-01 - 1	DL-SB02A	996	0.2887	-7471A	PPB	18 Jul 2017 10:56:32	U	SMPL
CCV32 - 1	CCV32	20313	5.3502	-7471A	PPB	18 Jul 2017 10:58:48	U	SMPL
CCB32 - 1	CCB32	-177	-0.0187	-7471A	PPB	18 Jul 2017 11:01:05	U	SMPL
I4203-02 - 1	DL-SB02C	1808	0.5014	-7471A	PPB	18 Jul 2017 11:03:22	U	SMPL
I4203-03 - 1	NB-SB02A	1182	0.3374	-7471A	PPB	18 Jul 2017 11:05:38	U	SMPL
I4203-04 - 1	NB-SB02B	60	0.0434	-7471A	PPB	18 Jul 2017 11:07:55	U	SMPL
I4203-05 - 1	NB-SB04A	10270	2.7187	-7471A	PPB	18 Jul 2017 11:10:11	U	SMPL
I4203-05DUP - 1	NB-SB04ADUP	10115	2.6781	-7471A	PPB	18 Jul 2017 11:12:28	U	SMPL
I4203-06 - 1	I4203-05MS	23251	6.1200	-7471A	PPB	18 Jul 2017 11:14:45	U	SMPL
I4203-07 - 1	I4203-05MSD	24833	6.5345	-7471A	PPB	18 Jul 2017 11:17:02	U	SMPL
I4203-08 - 1	NB-SB04B	-51	0.0143	-7471A	PPB	18 Jul 2017 11:19:21	U	SMPL
I4203-09 - 1	NB-SB01A	8682	2.3026	-7471A	PPB	18 Jul 2017 11:21:39	U	SMPL
I4203-10 - 1	NB-SB01B	51	0.0411	-7471A	PPB	18 Jul 2017 11:23:56	U	SMPL
CCV33 - 1	CCV33	20458	5.3882	-7471A	PPB	18 Jul 2017 11:26:12	U	SMPL
CCB33 - 1	CCB33	-141	-0.0092	-7471A	PPB	18 Jul 2017 11:28:28	U	SMPL
I4203-11 - 1	NB-SB03A	258480	67.7556	-7471A	PPB	18 Jul 2017 11:30:45	U	SMPL
I4203-12 - 1	NB-SB03B	155	0.0683	-7471A	PPB	18 Jul 2017 11:38:21	U	SMPL
I4223-01 - 1	WC071417-1N	174	0.0733	-7471A	PPB	18 Jul 2017 11:40:37	U	SMPL
I4223-04 - 1	WC071417-1NE	35	0.0369	-7471A	PPB	18 Jul 2017 11:42:54	U	SMPL
I4223-07 - 1	WC071417-1E	83	0.0495	-7471A	PPB	18 Jul 2017 11:45:10	U	SMPL
I4223-10 - 1	WC071417-1SE	200	0.0801	-7471A	PPB	18 Jul 2017 11:47:27	U	SMPL
I4187-09LX5 - 1		1478	0.4150	-7471A	PPB	18 Jul 2017 11:49:45	U	SMPL
I4187-09A - 1		23008	6.0563	-7471A	PPB	18 Jul 2017 11:52:02	U	SMPL
I4203-05LX5 - 1		1923	0.5316	-7471A	PPB	18 Jul 2017 11:54:20	U	SMPL
I4203-05A - 1		26048	6.8529	-7471A	PPB	18 Jul 2017 11:56:37	U	SMPL
CCV34 - 1	CCV34	19810	5.2184	-7471A	PPB	18 Jul 2017 11:58:54	U	SMPL
CCB34 - 1	CCB34	-206	-0.0263	-7471A	PPB	18 Jul 2017 12:01:12	U	SMPL
I4187-04DLX5 - 1		10825	2.8641	-7471A	PPB	18 Jul 2017 12:03:29	U	SMPL
I4203-11DLX10 - 1		28185	7.4128	-7471A	PPB	18 Jul 2017 12:05:45	U	SMPL
CCV35 - 1	CCV35	20138	5.3043	-7471A	PPB	18 Jul 2017 12:08:02	U	SMPL
CCB35 - 1	CCB35	-178	-0.0189	-7471A	PPB	18 Jul 2017 12:10:20	U	SMPL