



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88760

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/17/2017	16:30
2	CAL2	1	Water			7.01	07/17/2017	16:31
3	CAL3	1	Water			10.02	07/17/2017	16:32
4	ICV	1	Water			7.00	07/17/2017	16:34
5	CCV1	1	Water			2.01	07/17/2017	16:35
6	I4234-03	1	Solid	20.03	20	9.86	07/17/2017	16:36
7	I4234-05	1	Solid	20.05	20	7.76	07/17/2017	16:37
8	I4243-07	1	Solid	20.04	20	8.61	07/17/2017	16:39
9	I4243-08	1	Solid	20.02	20	7.70	07/17/2017	16:40
10	I4245-03	1	Solid	20.04	20	9.11	07/17/2017	16:41
11	I4245-06	1	Solid	20.03	20	7.67	07/17/2017	16:42
12	I4245-09	1	Solid	20.04	20	7.96	07/17/2017	16:43
13	I4245-12	1	Solid	20.02	20	8.53	07/17/2017	16:45
14	I4245-15	1	Solid	20.01	20	8.99	07/17/2017	16:46
15	I4245-15DUP	1	Solid	20.03	20	9.00	07/17/2017	16:47
16	CCV2	1	Water			12.01	07/17/2017	16:48

LB88760

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4234 WorkList ID : 100749 Date : 7/17/2017 3:54:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4234-03	Corrosivity	Cool 4 deg C	DIST01	G32	M-COMB-07	07/17/2017	9045C
	Solid	I4234-05	Corrosivity	Cool 4 deg C	DIST01	G32	M-COMB-02	07/17/2017	9045C
	Solid	I4243-07	Corrosivity	Cool 4 deg C	TECT01	F31	TP-CF	07/17/2017	9045C
	Solid	I4243-08	Corrosivity	Cool 4 deg C	TECT01	F31	TP-CN	07/17/2017	9045C
07/20/2017	Solid	I4245-03	Corrosivity	Cool 4 deg C	PSEG05	G32	PL-01-071717-A	07/17/2017	9045C
07/20/2017	Solid	I4245-06	Corrosivity	Cool 4 deg C	PSEG05	G32	PL-01-071717-B	07/17/2017	9045C
07/20/2017	Solid	I4245-09	Corrosivity	Cool 4 deg C	PSEG05	G32	PL-01-071717-C	07/17/2017	9045C
07/20/2017	Solid	I4245-12	Corrosivity	Cool 4 deg C	PSEG05	G32	PL-01-071717-D	07/17/2017	9045C
07/20/2017	Solid	I4245-15	Corrosivity	Cool 4 deg C	PSEG05	G32	PL-01-071717-E	07/17/2017	9045C

Date/Time 07-17-17 6:45 PM
Received by: JP
Relinquished by: JP

Date/Time 07-17-17 6:45 PM
Received by: JP
Relinquished by: JP