



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88762

Temp.: 22.4 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/17/2017	18:10
2	CAL2	1	Water			7.00	07/17/2017	18:11
3	CAL3	1	Water			10.03	07/17/2017	18:12
4	ICV	1	Water			7.01	07/17/2017	18:15
5	CCV1	1	Water			2.01	07/17/2017	18:16
6	I4246-08	1	Water			6.78	07/17/2017	18:17
7	I4246-08DUP	1	Water			6.79	07/17/2017	18:18
8	I4246-11	1	Water			7.11	07/17/2017	18:20
9	CCV2	1	Water			12.01	07/17/2017	18:21

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4246

WorkList ID : 100753

Date : 7/17/2017 6:01:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/17/2017	Water	I4246-08 - ☒	pH	Cool 4 deg C	PSEG01	G31	DRUM-1-2-3-COMP	07/17/2017	9045C
07/17/2017	Water	I4246-11 - ☒	pH	Cool 4 deg C	PSEG01	G31	DRUM-4-5-6-7-COMP	07/17/2017	9045C

Date/Time 07-17-17 6:00 PM
 Received by: JP
 Relinquished by: JP

Date/Time 07-17-17 7:00 PM
 Received by: JP
 Relinquished by: JP