

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 07/19/2017

DATE & TIME IN: 07/18/2017 17:00

OVEN TEMP: 98 °C

Run Number: LB88765

DATE & TIME OUT: 07/19/2017 12:50

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB88765BL	LB88765BL	87.5631	87.5631	100	87.5631	87.5631	87.5631	0	0
2	I4234-01	M-COMB-01	95.5019	95.5019	100	95.5405	95.5405	95.5405	0.0386	386
3	I4234-02	M-COMB-07	83.565	83.565	100	83.6014	83.6014	83.6014	0.0364	364
4	I4234-04	M-COMB-02	79.9765	79.9765	100	80.0169	80.0169	80.0169	0.0404	404
5	I4234-04DUP	M-COMB-02DUP	86.5993	86.5993	100	86.64	86.64	86.64	0.0407	407

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS 14234

WorkList ID : 100786

Date : 7/18/2017 11:36:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	14234-01-01	TS	Cool 4 deg C	DIST01	G32	M-COMB-01	07/17/2017	SM2540 B
	Water	14234-02-01	TS	Cool 4 deg C	DIST01	G32	M-COMB-07	07/17/2017	SM2540 B
	Water	14234-04-01	TS	Cool 4 deg C	DIST01	G32	M-COMB-02	07/17/2017	SM2540 B

Date/Time 07-18-17 11:36 PM
 Received by: 30
 Relinquished by:

Date/Time 07-18-17 5:25 PM
 Received by:
 Relinquished by: