


ISM02.4

LB 88770
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37715



A= 0.0000e+000

B= 2.6435e-004

C= 4.1022e-003

Rho= 0.9999898

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	-0.006	-0.006	-37	0.000	-37					-4
0.2	0.200	0.193	-0.007	715	0.0 %	715					0
2.5	2.500	2.496	-0.004	9428	0.0 %	9428					0
5.0	5.000	5.019	0.019	18971	0.0 %	18971					0
7.5	7.500	7.523	0.023	28443	0.0 %	28443					0
10.0	10.000	9.974	-0.026	37715	0.0 %	37715					0

ns
2/18/17

LB88770
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-37	-		0 ISM02.4	PPB	7/18/2017 14:48	Std
	0.2 S0.2	715	-		0.2 ISM02.4	PPB	7/18/2017 14:50	Std
	2.5 S2.5	9428	-		2.5 ISM02.4	PPB	7/18/2017 14:52	Std
	5 S5	18971	-		5 ISM02.4	PPB	7/18/2017 14:55	Std
	7.5 S7.5	28443	-		7.5 ISM02.4	PPB	7/18/2017 14:57	Std
	10 S10	37715	-		10 ISM02.4	PPB	7/18/2017 14:59	Std
ICV	ICV	14641	3.8745 -		ISM02.4	PPB	7/18/2017 15:02	SMPL
ICB	ICB	-198	-0.0482 -		ISM02.4	PPB	7/18/2017 15:04	SMPL
CCV036	CCV036	18908	5.0024 -		ISM02.4	PPB	7/18/2017 15:07	SMPL
CCB036	CCB036	-200	-0.0488 -		ISM02.4	PPB	7/18/2017 15:09	SMPL
PB100737BL	PBW010	-62	-0.0123 -		ISM02.4	PPB	7/18/2017 15:11	SMPL
PB100737TB	LEB010	-79	-0.0168 -		ISM02.4	PPB	7/18/2017 15:13	SMPL
I4208-01	MBC886	243	0.0683 -		ISM02.4	PPB	7/18/2017 15:16	SMPL
I4208-02	MBC887	132	0.039 -		ISM02.4	PPB	7/18/2017 15:18	SMPL
I4208-03	MBC888	136	0.0401 -		ISM02.4	PPB	7/18/2017 15:20	SMPL
I4208-04	MBC889	86	0.0268 -		ISM02.4	PPB	7/18/2017 15:22	SMPL
I4208-05	MBC889D	85	0.0266 -		ISM02.4	PPB	7/18/2017 15:25	SMPL
I4208-06	MBC889S	3991	1.0591 -		ISM02.4	PPB	7/18/2017 15:27	SMPL
CCV037	CCV037	19006	5.0283 -		ISM02.4	PPB	7/18/2017 15:29	SMPL
CCB037	CCB037	-174	-0.0419 -		ISM02.4	PPB	7/18/2017 15:32	SMPL