


7470A

LB88791
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32625



A= 0.0000e+000

B= 3.0932e-004

C= -6.4600e-002

Rho= 0.9999032

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.042	-0.042	72	0.000	72				
0.2	0.200	0.196	-0.004	843	0.0 %	843				
2.5	2.500	2.566	0.066	8505	0.0 %	8505				
5.0	5.000	5.039	0.039	16500	0.0 %	16500				
7.5	7.500	7.414	-0.086	24178	0.0 %	24178				
10.0	10.000	10.027	0.027	32625	0.0 %	32625				

LB88791

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jul 2017 10:17:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	72	-	0.0000	7470A	PPB	19 Jul 2017 10:22:46	S	Std
0.2 - 1	50.2	843	-	0.2000	7470A	PPB	19 Jul 2017 10:24:55	S	Std
2.5 - 1	52.5	8505	-	2.5000	7470A	PPB	19 Jul 2017 10:27:03	S	Std
5.0 - 1	55	16500	-	5.0000	7470A	PPB	19 Jul 2017 10:29:12	S	Std
7.5 - 1	57.5	24178	-	7.5000	7470A	PPB	19 Jul 2017 10:31:19	S	Std
10.0 - 1	510	32625	-	10.0000	7470A	PPB	19 Jul 2017 10:33:28	S	Std
ICV64 - 1	ICV64	12911	3.9290	-7470A	PPB	19 Jul 2017 10:36:08	U	SMPL	
ICB64 - 1	ICB64	-53	-0.0810	-7470A	PPB	19 Jul 2017 10:38:14	U	SMPL	
CCV38 - 1	CCV38	16824	5.1393	-7470A	PPB	19 Jul 2017 10:40:21	U	SMPL	
CCB38 - 1	CCB38	19	-0.0587	-7470A	PPB	19 Jul 2017 10:42:28	U	SMPL	
CRA - 1	CRA	881	0.2079	-7470A	PPB	19 Jul 2017 10:44:34	U	SMPL	
HighStd - 1	HighStd	33211	10.2081	-7470A	PPB	19 Jul 2017 10:46:41	U	SMPL	
ChkStd - 1	ChkStd	22421	6.8706	-7470A	PPB	19 Jul 2017 10:48:48	U	SMPL	
PB100747BL - 1	PBW	-81	-0.0897	-7470A	PPB	19 Jul 2017 10:53:20	U	SMPL	
PB100747BS - 1	LCSW	11590	3.5204	-7470A	PPB	19 Jul 2017 10:55:27	U	SMPL	
PB100747TB - 1	PB100700TB	-1	-0.0649	-7470A	PPB	19 Jul 2017 11:00:20	U	SMPL	
I4234-03 - 1	M-COMB-07	71	-0.0426	-7470A	PPB	19 Jul 2017 11:02:27	U	SMPL	
I4234-05 - 1	M-COMB-02	296	0.0270	-7470A	PPB	19 Jul 2017 11:04:35	U	SMPL	
I4243-07 - 1	TP-CF	-5	-0.0661	-7470A	PPB	19 Jul 2017 11:06:44	U	SMPL	
I4243-08 - 1	TP-CN	29	-0.0556	-7470A	PPB	19 Jul 2017 11:08:50	U	SMPL	
CCV39 - 1	CCV39	16500	5.0391	-7470A	PPB	19 Jul 2017 11:10:57	U	SMPL	
CCB39 - 1	CCB39	-55	-0.0816	-7470A	PPB	19 Jul 2017 11:13:04	U	SMPL	
I4245-03 - 1	PL-01-071717-A	4	-0.0634	-7470A	PPB	19 Jul 2017 11:15:11	U	SMPL	
I4245-06 - 1	PL-01-071717-B	81	-0.0395	-7470A	PPB	19 Jul 2017 11:17:18	U	SMPL	
I4245-09 - 1	PL-01-071717-C	104	-0.0324	-7470A	PPB	19 Jul 2017 11:19:25	U	SMPL	
I4245-12 - 1	PL-01-071717-D	113	-0.0296	-7470A	PPB	19 Jul 2017 11:21:32	U	SMPL	
I4245-15 - 1	PL-01-071717-E	132	-0.0238	-7470A	PPB	19 Jul 2017 11:23:39	U	SMPL	
I4245-15DUP - 1	PL-01-071717-EDUP	101	-0.0334	-7470A	PPB	19 Jul 2017 11:25:47	U	SMPL	
I4245-15MS - 1	PL-01-071717-EMS	12527	3.8102	-7470A	PPB	19 Jul 2017 11:27:55	U	SMPL	
I4245-15MSD - 1	PL-01-071717-EMSD	12503	3.8028	-7470A	PPB	19 Jul 2017 11:30:03	U	SMPL	
I4246-09 - 1	DRUM-1-2-3-COMP	-25	-0.0723	-7470A	PPB	19 Jul 2017 11:32:11	U	SMPL	
I4246-12 - 1	DRUM-4-5-6-7-COMP	5	-0.0631	-7470A	PPB	19 Jul 2017 11:34:18	U	SMPL	
CCV40 - 1	CCV40	16716	5.1059	-7470A	PPB	19 Jul 2017 11:36:25	U	SMPL	
CCB40 - 1	CCB40	-65	-0.0847	-7470A	PPB	19 Jul 2017 11:38:32	U	SMPL	
I4246-15 - 1	DRUM-8-9-COMP-STONE	46	-0.0504	-7470A	PPB	19 Jul 2017 11:40:39	U	SMPL	
PB100748BL - 1	PBW	27	-0.0562	-7470A	PPB	19 Jul 2017 11:42:46	U	SMPL	
PB100748BS - 1	LCSW	11884	3.6113	-7470A	PPB	19 Jul 2017 11:44:53	U	SMPL	
I4246-08 - 1	DRUM-1-2-3-COMP	9115	2.7548	-7470A	PPB	19 Jul 2017 11:47:00	U	SMPL	
I4246-08DUP - 1	DRUM-1-2-3-COMP-DUP	8508	2.5671	-7470A	PPB	19 Jul 2017 11:49:07	U	SMPL	
I4246-08MS - 1	DRUM-1-2-3-COMPMS	21862	6.6977	-7470A	PPB	19 Jul 2017 11:51:15	U	SMPL	
I4246-08MSD - 1	DRUM-1-2-3-COMPMSD	22193	6.8001	-7470A	PPB	19 Jul 2017 11:53:23	U	SMPL	
I4246-11 - 1	DRUM-4-5-6-7-COMP	9252	2.7972	-7470A	PPB	19 Jul 2017 11:55:32	U	SMPL	
CCV41 - 1	CCV41	16665	5.0902	-7470A	PPB	19 Jul 2017 11:57:42	U	SMPL	
CCB41 - 1	CCB41	23	-0.0575	-7470A	PPB	19 Jul 2017 12:03:23	U	SMPL	
I4245-15LX5 - 1		79	-0.0402	-7470A	PPB	19 Jul 2017 12:05:30	U	SMPL	
I4245-15A - 1		13127	3.9958	-7470A	PPB	19 Jul 2017 12:07:37	U	SMPL	
I4246-08LX5 - 1		1562	0.4186	-7470A	PPB	19 Jul 2017 12:09:44	U	SMPL	
I4246-08A - 1		21845	6.6924	-7470A	PPB	19 Jul 2017 12:11:51	U	SMPL	
CCV42 - 1	CCV42	16503	5.0401	-7470A	PPB	19 Jul 2017 12:13:58	U	SMPL	
CCB42 - 1	CCB42	34	-0.0541	-7470A	PPB	19 Jul 2017 12:18:26	U	SMPL	