



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB88793

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP56166
COD calibration std. 100 ppm	WP56164
COD calibration std. 10 ppm	WP56162
COD calibration std. 150 ppm	WP56165
COD calibration std. 50 ppm	WP56163
COD calibration std. 0 ppm	WP56161
COD Digestion Vials Low Level 0-150Mg/L	W2235
COD ICV-LCS std, 50ppm	WP56500
COD CCV std, 50ppm	WP56501

Temp In (C): <u>150</u>	Date In: <u>07/19/2017</u>	Time In: <u>09:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>07/19/2017</u>	Time Out: <u>11:02</u>

Intercept: 0.2985

Slope: 1.0016

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/05/2017	16:00
2	CAL2	10	1	Water	10.000	07/05/2017	16:00
3	CAL3	50	1	Water	51.000	07/05/2017	16:01
4	CAL4	100	1	Water	101.000	07/05/2017	16:01
5	CAL5	150	1	Water	150.000	07/05/2017	16:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB88793

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.622	07/05/2017	16:02
2	ICB				1	Water	0.000	-0.298	07/05/2017	16:03
3	CCV1	50			1	Water	52.000	51.619	07/19/2017	12:00
4	CCB1				1	Water	0.000	-0.298	07/19/2017	12:00
5	LB88793BL				1	Water	0.000	-0.298	07/19/2017	12:01
6	LB88793BS	50			1	Water	51.000	50.621	07/19/2017	12:01
7	I4246-08				5	Water	40.000	39.638	07/19/2017	12:02
8	I4246-11				5	Water	51.000	50.621	07/19/2017	12:02
9	I4246-11DUP				5	Water	52.000	51.619	07/19/2017	12:03
10	I4246-11MS	50			5	Water	60.000	59.606	07/19/2017	12:03
11	I4246-11MSD	50			5	Water	61.000	60.605	07/19/2017	12:04
12	CCV2	50			1	Water	52.000	51.619	07/19/2017	12:04
13	CCB2				1	Water	0.000	-0.298	07/19/2017	12:05

LB 88793

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD071817

WorkList ID : 100787

Date : 7/18/2017 11:45:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2017	Water	I4246-08	COD	Conc H2SO4 to	PSEG01	G31	DRUM-1-2-3-COMP	07/17/2017	SM5220 D
07/20/2017	Water	I4246-11	COD	Conc H2SO4 to	PSEG01	G31	DRUM-4-5-6-7-COMP	07/17/2017	SM5220 D

Date/Time 07/19/17 8:50
Received by: LB
Relinquished by: CP

Date/Time 07/19/17 10:00
Received by: CP
Relinquished by: LB