


245.1

LB88797
INSTRUMENT ID: CV1

Linear

 μ Abs.:

28559



A= 0.0000e+000

B= 3.5013e-004

C= -4.6517e-002

Rho= 0.9999640

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	53	0.000	53				
0.2	0.200	0.186	-0.014	664	0.0 %	664				
2.5	2.500	2.525	0.025	7344	0.0 %	7344				
5.0	5.000	5.030	0.030	14500	0.0 %	14500				
7.5	7.500	7.534	0.034	21650	0.0 %	21650				
10.0	10.000	9.953	-0.047	28559	0.0 %	28559				

LB88797 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Jul 2017 12:28:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000	245.1	PPB	19 Jul 2017 12:30:42	S	Std
0.2 - 1	50.2	664	-	0.2000	245.1	PPB	19 Jul 2017 12:32:50	S	Std
2.5 - 1	52.5	7344	-	2.5000	245.1	PPB	19 Jul 2017 12:34:58	S	Std
5.0 - 1	55	14500	-	5.0000	245.1	PPB	19 Jul 2017 12:37:07	S	Std
7.5 - 1	57.5	21650	-	7.5000	245.1	PPB	19 Jul 2017 12:39:14	S	Std
10.0 - 1	510	28559	-	10.0000	245.1	PPB	19 Jul 2017 12:41:22	S	Std
ICV65 - 1	ICV65	11386	3.9401	-245.1		PPB	19 Jul 2017 12:44:01	U	SMPL
ICB65 - 1	ICB65	-47	-0.0630	-245.1		PPB	19 Jul 2017 12:46:07	U	SMPL
CCV43 - 1	CCV43	14592	5.0626	-245.1		PPB	19 Jul 2017 12:48:14	U	SMPL
CCB43 - 1	CCB43	-57	-0.0665	-245.1		PPB	19 Jul 2017 12:50:21	U	SMPL
CRA - 1	CRA	659	0.1842	-245.1		PPB	19 Jul 2017 12:52:29	U	SMPL
HighStd - 1	HighStd	28828	10.0471	-245.1		PPB	19 Jul 2017 12:54:36	U	SMPL
ChkStd - 1	ChkStd	19838	6.8994	-245.1		PPB	19 Jul 2017 12:56:43	U	SMPL
PB100749BL - 1	PBW	10	-0.0430	-245.1		PPB	19 Jul 2017 13:01:17	U	SMPL
PB100749BS - 1	LCSW	12007	4.1575	-245.1		PPB	19 Jul 2017 13:03:24	U	SMPL
I4234-01 - 1	M-COMB-01	111	-0.0077	-245.1		PPB	19 Jul 2017 13:05:32	U	SMPL
I4234-01DUP - 1	M-COMB-01DUP	225	0.0323	-245.1		PPB	19 Jul 2017 13:07:40	U	SMPL
I4234-01MS - 1	M-COMB-01MS	12379	4.2878	-245.1		PPB	19 Jul 2017 13:09:48	U	SMPL
I4234-01MSD - 1	M-COMB-01MSD	12664	4.3876	-245.1		PPB	19 Jul 2017 13:11:57	U	SMPL
I4234-02 - 1	M-COMB-07	106	-0.0094	-245.1		PPB	19 Jul 2017 13:14:04	U	SMPL
CCV44 - 1	CCV44	14615	5.0707	-245.1		PPB	19 Jul 2017 13:16:10	U	SMPL
CCB44 - 1	CCB44	-44	-0.0619	-245.1		PPB	19 Jul 2017 13:18:17	U	SMPL
I4234-04 - 1	M-COMB-02	127	-0.0021	-245.1		PPB	19 Jul 2017 13:20:25	U	SMPL
I4274-01 - 1	M-COMB-3	117	-0.0056	-245.1		PPB	19 Jul 2017 13:22:31	U	SMPL
I4274-03 - 1	M-COMB-4	111	-0.0077	-245.1		PPB	19 Jul 2017 13:24:38	U	SMPL
I4274-04 - 1	M-COMB-05	220	0.0305	-245.1		PPB	19 Jul 2017 13:26:45	U	SMPL
I4274-06 - 1	M-COMB-6	81	-0.0182	-245.1		PPB	19 Jul 2017 13:28:53	U	SMPL
I4234-01LX5 - 1		78	-0.0192	-245.1		PPB	19 Jul 2017 13:31:01	U	SMPL
I4234-01A - 1		12708	4.4030	-245.1		PPB	19 Jul 2017 13:33:09	U	SMPL
CCV45 - 1	CCV45	14384	4.9898	-245.1		PPB	19 Jul 2017 13:35:17	U	SMPL
CCB45 - 1	CCB45	-56	-0.0661	-245.1		PPB	19 Jul 2017 13:37:26	U	SMPL