

7471A


LB88800

INSTRUMENT ID: CY2

Linear

 μ Abs.:

39822



A= 0.0000e+000

B= 2.5334e-004

C= 3.0327e-003

Rho= 0.9998187

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	18	0.000	18				
0.2	0.200	0.208	0.008	810	0.0 %	810				
2.5	2.500	2.520	0.020	9937	0.0 %	9937				
5.0	5.000	5.016	0.016	19786	0.0 %	19786				
7.5	7.500	7.357	-0.143	29028	0.0 %	29028	0			
10.0	10.000	10.091	0.091	39822	0.0 %	39822				

LB88800 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 19 Jul 2017 14:30:21

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	7471A	PPB	19 Jul 2017 14:32:01	S	Std
0.2 - 1	50.2	810	-	0.2000	7471A	PPB	19 Jul 2017 14:34:19	S	Std
2.5 - 1	52.5	9937	-	2.5000	7471A	PPB	19 Jul 2017 14:36:37	S	Std
5.0 - 1	55	19786	-	5.0000	7471A	PPB	19 Jul 2017 14:38:54	S	Std
7.5 - 1	57.5	29028	-	7.5000	7471A	PPB	19 Jul 2017 14:41:11	S	Std
10.0 - 1	510	39822	-	10.0000	7471A	PPB	19 Jul 2017 14:48:38	S	Std
ICV66 - 1	ICV66	14226	3.6070	-	7471A	PPB	19 Jul 2017 14:51:22	U	SMPL
ICB66 - 1	ICB66	-119	-0.0271	-	7471A	PPB	19 Jul 2017 14:53:37	U	SMPL
CCV46 - 1	CCV46	18715	4.7442	-	7471A	PPB	19 Jul 2017 14:55:53	U	SMPL
CCB46 - 1	CCB46	-143	-0.0332	-	7471A	PPB	19 Jul 2017 14:58:09	U	SMPL
CRA - 1	CRA	770	0.1981	-	7471A	PPB	19 Jul 2017 15:00:25	U	SMPL
HighStd - 1	HighStd	39821	10.0911	-	7471A	PPB	19 Jul 2017 15:02:41	U	SMPL
ChkStd - 1	ChkStd	26823	6.7983	-	7471A	PPB	19 Jul 2017 15:05:44	U	SMPL
PB100751BL - 1	PBS	-181	-0.0428	-	7471A	PPB	19 Jul 2017 15:08:00	U	SMPL
PB100751BS - 1	LCSS	14703	3.7278	-	7471A	PPB	19 Jul 2017 15:10:19	U	SMPL
I4146-05 - 1	FK-SF-03-001-C	1873	0.4775	-	7471A	PPB	19 Jul 2017 15:12:36	U	SMPL
I4146-05 - 1	FK-SF-03-001-C	1931	0.4922	-	7471A	PPB	19 Jul 2017 15:15:11	U	SMPL
I4243-03 - 1	TP-CF	7964	2.0206	-	7471A	PPB	19 Jul 2017 15:20:10	U	SMPL
I4243-04 - 1	TP-CN	1098	0.2812	-	7471A	PPB	19 Jul 2017 15:22:30	U	SMPL
I4245-01 - 1	PL-01-071717-A	688	0.1773	-	7471A	PPB	19 Jul 2017 15:27:04	U	SMPL
CCV47 - 1	CCV47	19489	4.9403	-	7471A	PPB	19 Jul 2017 15:29:20	U	SMPL
CCB47 - 1	CCB47	-105	-0.0236	-	7471A	PPB	19 Jul 2017 15:31:36	U	SMPL
I4245-04 - 1	PL-01-071717-B	839	0.2156	-	7471A	PPB	19 Jul 2017 15:33:53	U	SMPL
I4245-04DUP - 1	PL-01-071717-BDUP	838	0.2153	-	7471A	PPB	19 Jul 2017 15:36:08	U	SMPL
I4245-04MS - 1	PL-01-071717-BMS	15725	3.9867	-	7471A	PPB	19 Jul 2017 15:38:24	U	SMPL
I4245-04MSD - 1	PL-01-071717-BMSD	15724	3.9865	-	7471A	PPB	19 Jul 2017 15:40:40	U	SMPL
I4245-07 - 1	PL-01-071717-C	588	0.1520	-	7471A	PPB	19 Jul 2017 15:42:57	U	SMPL
I4245-10 - 1	PL-01-071717-D	573	0.1482	-	7471A	PPB	19 Jul 2017 15:45:14	U	SMPL
I4245-13 - 1	PL-01-071717-E	377	0.0985	-	7471A	PPB	19 Jul 2017 15:47:31	U	SMPL
I4246-14 - 1	DRUM-8-9-COMP-STONE	22	0.0086	-	7471A	PPB	19 Jul 2017 15:49:48	U	SMPL
I4248-01 - 1	TP-3-COMP-(0-8)	23152	5.8683	-	7471A	PPB	19 Jul 2017 15:52:06	U	SMPL
I4248-05 - 1	TP-1-COMP-(0-8)	34464	8.7340	-	7471A	PPB	19 Jul 2017 15:54:22	U	SMPL
CCV48 - 1	CCV48	19283	4.8881	-	7471A	PPB	19 Jul 2017 15:56:39	U	SMPL
CCB48 - 1	CCB48	-139	-0.0322	-	7471A	PPB	19 Jul 2017 15:58:57	U	SMPL
I4250-01 - 1	ROLL-OFF-4611-COMP	11701	2.9673	-	7471A	PPB	19 Jul 2017 16:01:14	U	SMPL
PB100752BL - 1	PBS	0	0.0030	-	7471A	PPB	19 Jul 2017 16:03:30	U	SMPL
PB100752BS - 1	LCSS	15024	3.8091	-	7471A	PPB	19 Jul 2017 16:05:47	U	SMPL
I4253-01 - 1	E2-O-P-6-7-BASE	153288	38.8363	-	7471A	PPB	19 Jul 2017 16:08:03	U	SMPL
I4253-02 - 1	E2-O-P-6-7-WSW	170794	43.2712	-	7471A	PPB	19 Jul 2017 16:10:20	U	SMPL
I4253-03 - 1	E2-O-P-6-7-SSW	120461	30.5201	-	7471A	PPB	19 Jul 2017 16:12:43	U	SMPL
I4253-04 - 1	E2-O-P-6-7-NSW	3600	0.9150	-	7471A	PPB	19 Jul 2017 16:15:08	U	SMPL
I4253-05 - 1	E2-06-BASE	32061	8.1252	-	7471A	PPB	19 Jul 2017 16:17:29	U	SMPL
I4253-06 - 1	E2-07-WSW	12588	3.1920	-	7471A	PPB	19 Jul 2017 16:19:47	U	SMPL
I4253-07 - 1	E2-07-BASE	155272	39.3389	-	7471A	PPB	19 Jul 2017 16:22:05	U	SMPL
CCV49 - 1	CCV49	18691	4.7381	-	7471A	PPB	19 Jul 2017 16:24:21	U	SMPL
CCB49 - 1	CCB49	-238	-0.0573	-	7471A	PPB	19 Jul 2017 16:26:46	U	SMPL
I4254-01 - 1	72-12019/72-11985/72-12016-COMP	1246	0.3187	-	7471A	PPB	19 Jul 2017 16:29:03	U	SMPL
I4258-01 - 1	OK-02-071817-A	261	0.0692	-	7471A	PPB	19 Jul 2017 16:31:20	U	SMPL
I4258-04 - 1	OK-02-071817-B	1398	0.3572	-	7471A	PPB	19 Jul 2017 16:33:36	U	SMPL
I4258-07 - 1	OK-02-071817-C	1683	0.4294	-	7471A	PPB	19 Jul 2017 16:35:53	U	SMPL
I4265-01 - 1	SU-03-071817	684	0.1763	-	7471A	PPB	19 Jul 2017 16:38:09	U	SMPL
I4265-01DUP - 1	SU-03-071817DUP	617	0.1593	-	7471A	PPB	19 Jul 2017 16:40:26	U	SMPL
I4265-01MS - 1	SU-03-071817MS	15547	3.9416	-	7471A	PPB	19 Jul 2017 16:42:44	U	SMPL
I4265-01MSD - 1	SU-03-071817MSD	15204	3.8547	-	7471A	PPB	19 Jul 2017 16:45:01	U	SMPL
I4272-01 - 1	HDD-P1LWC2	-23	-0.0028	-	7471A	PPB	19 Jul 2017 16:47:18	U	SMPL
I4276-01 - 1	TP-2-COMP-(0-8)	137789	34.9099	-	7471A	PPB	19 Jul 2017 16:49:34	U	SMPL
CCV50 - 1	CCV50	18611	4.7179	-	7471A	PPB	19 Jul 2017 16:51:51	U	SMPL
CCB50 - 1	CCB50	-193	-0.0459	-	7471A	PPB	19 Jul 2017 16:54:12	U	SMPL
I4276-05 - 1	TP-6-COMP-(0-8)	122634	31.0706	-	7471A	PPB	19 Jul 2017 16:56:30	U	SMPL
I4280-01 - 1	CL-01-071817-A	123101	31.1889	-	7471A	PPB	19 Jul 2017 16:58:46	U	SMPL
I4280-04 - 1	CL-01-071817-B	3253	0.8271	-	7471A	PPB	19 Jul 2017 17:01:05	U	SMPL
I4280-07 - 1	CL-01-071817-C	1936	0.4935	-	7471A	PPB	19 Jul 2017 17:03:25	U	SMPL
I4280-10 - 1	CL-01-071817-D	1397	0.3569	-	7471A	PPB	19 Jul 2017 17:05:41	U	SMPL
I4245-04LX5 - 1		62	0.0187	-	7471A	PPB	19 Jul 2017 17:15:40	U	SMPL
I4245-04A - 1		16455	4.1717	-	7471A	PPB	19 Jul 2017 17:17:57	U	SMPL
I4265-01LX5 - 1		57	0.0175	-	7471A	PPB	19 Jul 2017 17:20:15	U	SMPL
I4265-01A - 1		16714	4.2373	-	7471A	PPB	19 Jul 2017 17:22:31	U	SMPL
CCV51 - 1	CCV51	19319	4.8972	-	7471A	PPB	19 Jul 2017 17:24:47	U	SMPL
CCB51 - 1	CCB51	-146	-0.0340	-	7471A	PPB	19 Jul 2017 17:27:04	U	SMPL
I4253-01DLX10 - 1		16677	4.2279	-	7471A	PPB	19 Jul 2017 17:29:21	U	SMPL
I4253-02DLX10 - 1		17990	4.5605	-	7471A	PPB	19 Jul 2017 17:31:38	U	SMPL
I4253-03DLX10 - 1		12561	3.1852	-	7471A	PPB	19 Jul 2017 17:33:55	U	SMPL
I4253-07DLX10 - 1		16407	4.1595	-	7471A	PPB	19 Jul 2017 17:36:11	U	SMPL
I4276-01DLX10 - 1		13585	3.4446	-	7471A	PPB	19 Jul 2017 17:38:28	U	SMPL
I4276-05DLX10 - 1		12171	3.0864	-	7471A	PPB	19 Jul 2017 17:40:45	U	SMPL
I4280-01DLX10 - 1		12441	3.1548	-	7471A	PPB	19 Jul 2017 17:43:02	U	SMPL
CCV52 - 1	CCV52	19375	4.9114	-	7471A	PPB	19 Jul 2017 17:45:20	U	SMPL
CCB52 - 1	CCB52	-113	-0.0256	-	7471A	PPB	19 Jul 2017 17:47:38	U	SMPL