

7470A


LB88830

INSTRUMENT ID: CVA

Linear

 μ Abs.:

34075



A= 0.0000e+000

B= 2.9236e-004

C= 3.0131e-002

Rho= 0.9999576

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	-26	0.000	-26				
0.2	0.200	0.206	0.006	603	0.0 %	603				
2.5	2.500	2.483	-0.017	8389	0.0 %	8389				
5.0	5.000	4.943	-0.057	16805	0.0 %	16805				
7.5	7.500	7.553	0.053	25731	0.0 %	25731				
10.0	10.000	9.992	-0.008	34075	0.0 %	34075				

Method: 7470A

Operator: Admin

LB88830 INSTRUMENT ID : CV1

Date of Analysis: 20 Jul 2017 11:26:42

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.0007470A	PPB		20 Jul 2017 11:56:15	S	Std
0.2 - 1	50.2	603	-	0.20007470A	PPB		20 Jul 2017 11:58:23	S	Std
2.5 - 1	52.5	8389	-	2.50007470A	PPB		20 Jul 2017 12:00:32	S	Std
5.0 - 1	55	16805	-	5.00007470A	PPB		20 Jul 2017 12:02:40	S	Std
7.5 - 1	57.5	25731	-	7.50007470A	PPB		20 Jul 2017 12:07:33	S	Std
10.0 - 1	510	34075	-	10.00007470A	PPB		20 Jul 2017 12:14:31	S	Std
ICV67 - 1	ICV67	12852	3.7875	-7470A	PPB		20 Jul 2017 12:17:08	U	SMPL
ICB67 - 1	ICB67	-159	-0.0164	-7470A	PPB		20 Jul 2017 12:19:14	U	SMPL
CCV53 - 1	CCV53	16694	4.9108	-7470A	PPB		20 Jul 2017 12:21:21	U	SMPL
CCB53 - 1	CCB53	-145	-0.0123	-7470A	PPB		20 Jul 2017 12:23:27	U	SMPL
CRA - 1	CRA	630	0.2143	-7470A	PPB		20 Jul 2017 12:25:35	U	SMPL
HighStd - 1	HighStd	34502	10.1171	-7470A	PPB		20 Jul 2017 12:27:41	U	SMPL
ChkStd - 1	ChkStd	22471	6.5997	-7470A	PPB		20 Jul 2017 12:29:48	U	SMPL
PB100775BL - 1	PBW	-188	-0.0248	-7470A	PPB		20 Jul 2017 12:31:56	U	SMPL
PB100775BS - 1	LCSW	12210	3.5998	-7470A	PPB		20 Jul 2017 12:34:05	U	SMPL
PB100775TB - 1	PB100743TB	-123	-0.0058	-7470A	PPB		20 Jul 2017 12:36:12	U	SMPL
I4248-03 - 1	TP-3-COMP-(0-8)	-36	0.0196	-7470A	PPB		20 Jul 2017 12:38:21	U	SMPL
I4248-07 - 1	TP-1-COMP-(0-8)	59	0.0474	-7470A	PPB		20 Jul 2017 12:40:29	U	SMPL
I4250-03 - 1	ROLL-OFF-4611-COMP	17	0.0351	-7470A	PPB		20 Jul 2017 12:42:37	U	SMPL
I4254-06 - 1	72-12019/72-11985/72-12016-COMP	-86	0.0050	-7470A	PPB		20 Jul 2017 12:44:44	U	SMPL
CCV54 - 1	CCV54	16897	4.9701	-7470A	PPB		20 Jul 2017 12:46:51	U	SMPL
CCB54 - 1	CCB54	-154	-0.0149	-7470A	PPB		20 Jul 2017 12:48:58	U	SMPL
I4258-03 - 1	OK-02-071817-A	-32	0.0208	-7470A	PPB		20 Jul 2017 12:51:05	U	SMPL
I4258-06 - 1	OK-02-071817-B	-57	0.0135	-7470A	PPB		20 Jul 2017 12:53:12	U	SMPL
I4258-09 - 1	OK-02-071817-C	-19	0.0246	-7470A	PPB		20 Jul 2017 12:55:18	U	SMPL
I4265-03 - 1	SU-03-071817	-54	0.0143	-7470A	PPB		20 Jul 2017 12:57:25	U	SMPL
I4272-02 - 1	HDD-P1LWC2	-59	0.0129	-7470A	PPB		20 Jul 2017 12:59:33	U	SMPL
I4274-02 - 1	M-COMB-4	126	0.0670	-7470A	PPB		20 Jul 2017 13:01:40	U	SMPL
I4274-05 - 1	M-COMB-05	-26	0.0225	-7470A	PPB		20 Jul 2017 13:03:48	U	SMPL
I4274-07 - 1	M-COMB-6	-33	0.0205	-7470A	PPB		20 Jul 2017 13:05:56	U	SMPL
I4276-03 - 1	TP-2-COMP-(0-8)	486	0.1722	-7470A	PPB		20 Jul 2017 13:08:05	U	SMPL
I4276-07 - 1	TP-6-COMP-(0-8)	-37	0.0193	-7470A	PPB		20 Jul 2017 13:10:12	U	SMPL
CCV55 - 1	CCV55	16657	4.9000	-7470A	PPB		20 Jul 2017 13:12:19	U	SMPL
CCB55 - 1	CCB55	-122	-0.0055	-7470A	PPB		20 Jul 2017 13:14:26	U	SMPL
I4276-07DUP - 1	TP-6-COMP-(0-8)DUP	-45	0.0170	-7470A	PPB		20 Jul 2017 13:16:33	U	SMPL
I4276-07MS - 1	TP-6-COMP-(0-8)MS	12324	3.6332	-7470A	PPB		20 Jul 2017 13:18:39	U	SMPL
I4276-07MSD - 1	TP-6-COMP-(0-8)MSD	12474	3.6770	-7470A	PPB		20 Jul 2017 13:20:46	U	SMPL
I4276-07LX5 - 1		-97	0.0018	-7470A	PPB		20 Jul 2017 13:22:53	U	SMPL
I4276-07A - 1		13589	4.0030	-7470A	PPB		20 Jul 2017 13:25:01	U	SMPL
CCV56 - 1	CCV56	16855	4.9578	-7470A	PPB		20 Jul 2017 13:27:09	U	SMPL
CCB56 - 1	CCB56	-93	0.0029	-7470A	PPB		20 Jul 2017 13:29:17	U	SMPL