



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88832

Temp.: 21.5 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/20/2017	13:35
2	CAL2	1	Water			7.01	07/20/2017	13:36
3	CAL3	1	Water			10.04	07/20/2017	13:37
4	ICV	1	Water			7.00	07/20/2017	13:38
5	CCV1	1	Water			2.02	07/20/2017	13:40
6	I4326-01	1	Water			3.97	07/20/2017	13:41
7	I4326-01DUP	1	Water			3.98	07/20/2017	13:42
8	I4326-02	1	Water			4.06	07/20/2017	13:44
9	I4326-03	1	Water			4.73	07/20/2017	13:46
10	CCV2	1	Water			12.01	07/20/2017	13:47

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-14274

WorkList ID : 100841

Date : 7/19/2017 2:32:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4274-01-G	TS	Cool 4 deg C	DIST01	D22	M-COMB-3	07/18/2017	SM2540 B
	Water	I4274-03-F	TS	Cool 4 deg C	DIST01	D22	M-COMB-4	07/18/2017	SM2540 B
	Water	I4274-04-F	TS	Cool 4 deg C	DIST01	D22	M-COMB-05	07/18/2017	SM2540 B
	Water	I4274-06-G	TS	Cool 4 deg C	DIST01	D22	M-COMB-6	07/18/2017	SM2540 B

Date/Time 07-19-17 15:00
 Received by: JD
 Relinquished by: CD

Date/Time 07-19-17 06:00
 Received by: JP
 Relinquished by: JP