



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88841

Temp.: 21.6 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/20/2017	16:20
2	CAL2	1	Water			7.00	07/20/2017	16:21
3	CAL3	1	Water			10.03	07/20/2017	16:22
4	ICV	1	Water			7.01	07/20/2017	16:23
5	CCV1	1	Water			2.01	07/20/2017	16:25
6	I4354-03	1	Water			7.48	07/20/2017	16:26
7	I4354-03DUP	1	Water			7.49	07/20/2017	16:30
8	CCV2	1	Water			12.01	07/20/2017	16:31

WORKLIST(Hardcopy Internal Chain)

2B88841

WorkList Name : PH I4354

WorkList ID : 100909

Date : 7/20/2017 4:06:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2017	Water	I4354-03-17	pH	Cool 4 deg C	PSEG01	H22	265434-265578-COMP	07/20/2017	9045C

Date/Time 07/20/17 4:06 PM
 Received by: JI
 Relinquished by: CA

Date/Time 07/20/17 3:00 PM
 Received by: CF
 Relinquished by: Jn