



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB88844

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	07/20/2017	10:05
2	I4234-01		1	>100.0	>212.0	07/20/2017	10:35
3	I4234-01DUP		1	>100.0	>212.0	07/20/2017	10:50
4	I4234-02		1	>100.0	>212.0	07/20/2017	11:10
5	I4234-04		1	>100.0	>212.0	07/20/2017	11:32
6	I4246-09		1	>100.0	>212.0	07/20/2017	11:55
7	I4246-12		1	>100.0	>212.0	07/20/2017	12:15
8	I4274-01		1	>100.0	>212.0	07/20/2017	12:40
9	I4274-03		1	>100.0	>212.0	07/20/2017	12:58
10	I4274-04		1	>100.0	>212.0	07/20/2017	13:20
11	I4274-06		1	>100.0	>212.0	07/20/2017	13:45
12	I4313-01		1	>100.0	>212.0	07/20/2017	14:10
13	I4313-02		1	>100.0	>212.0	07/20/2017	14:40
14	I4313-03		1	>100.0	>212.0	07/20/2017	15:05
15	I4354-04		1	>100.0	>212.0	07/20/2017	15:35

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 88844

Worklist Name : FP72017

Worklist ID : 100884

Date : 7/20/2017 9:34:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4234-01	Flash Point	Cool 4 deg C	DIST01	G32	M-COMB-01	07/17/2017	1010A
	Water	I4234-02	Flash Point	Cool 4 deg C	DIST01	G32	M-COMB-07	07/17/2017	1010A
	Water	I4234-04	Flash Point	Cool 4 deg C	DIST01	G32	M-COMB-02	07/17/2017	1010A
07/20/2017	Water	I4246-09	Flash Point	Cool 4 deg C	PSEG01	G31	DRUM-1-2-3-COMP	07/17/2017	1010A
07/20/2017	Water	I4246-12	Flash Point	Cool 4 deg C	PSEG01	G31	DRUM-4-5-6-7-COMP	07/17/2017	1010A
	Water	I4274-01	Flash Point	Cool 4 deg C	DIST01	D22	M-COMB-3	07/18/2017	1010A
	Water	I4274-03	Flash Point	Cool 4 deg C	DIST01	D22	M-COMB-4	07/18/2017	1010A
	Water	I4274-04	Flash Point	Cool 4 deg C	DIST01	D22	M-COMB-05	07/18/2017	1010A
	Water	I4274-06	Flash Point	Cool 4 deg C	DIST01	D22	M-COMB-6	07/18/2017	1010A
	Water	I4313-01	Flash Point	Cool 4 deg C	DIST01	G41	M-COMB-8	07/19/2017	1010A
	Water	I4313-02	Flash Point	Cool 4 deg C	DIST01	G41	M-COMB-9	07/19/2017	1010A
	Water	I4313-03	Flash Point	Cool 4 deg C	DIST01	G41	B-COMB-1	07/19/2017	1010A
07/25/2017	Water	I4354-04	Flash Point	Cool 4 deg C	PSEG01	H22	265434-265578-COMP	07/20/2017	1010A

Date/Time

07/20/17 9:40 AM

Received by:

KS

Relinquished by:

CS

Date/Time

07/20/17 15:40

Received by:

CS

Relinquished by:

KS