



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88846

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/20/2017	16:35
2	CAL2	1	Water			7.01	07/20/2017	16:36
3	CAL3	1	Water			10.02	07/20/2017	16:37
4	ICV	1	Water			7.00	07/20/2017	16:38
5	CCV1	1	Water			2.01	07/20/2017	16:40
6	I4347-03	1	Solid	20.02	20	8.45	07/20/2017	16:41
7	I4347-03DUP	1	Solid	20.04	20	8.46	07/20/2017	16:42
8	I4358-03	1	Solid	20.01	20	7.91	07/20/2017	16:43
9	I4363-03	1	Solid	20.04	20	7.90	07/20/2017	16:44
10	CCV2	1	Water			12.01	07/20/2017	16:45

WORKLIST(Hardcopy Internal Chain)

288846

WorkList Name : CORROSIVITY I4304

WorkList ID : 100901

Date : 7/20/2017 2:38:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4347-03	Corrosivity	Cool 4 deg C	WSP001	H31	B-2(6-10)COMP	07/19/2017	9045C
07/25/2017	Solid	I4358-03	Corrosivity	Cool 4 deg C	PSEG05	H23	TR-05-072017	07/20/2017	9045C
	Solid	I4363-03	Corrosivity	Cool 4 deg C	DIST01	H32	B-COMB-4	07/20/2017	9045C

Date/Time 07/20/17 2:40PM
 Received by: JH
 Relinquished by: CP

Date/Time 07/20/17 6:20PM
 Received by: CP
 Relinquished by: JO