

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 07/21/2017

DATE & TIME IN: 07/21/2017 09:10

OVEN TEMP: 104 °C

Run Number: LB88855

DATE & TIME OUT: 07/21/2017 13:15

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB88855BL	LB88855BL	1.4668	1.4668	100	1.4668	1.4668	1.4668	0	0
2	LB88855BS	LB88855BS	1.3428	1.3428	100	1.3973	1.3973	1.3973	0.0545	545
3	I4344-01	SPDES-1	1.4057	1.4057	900	1.406	1.406	1.406	0.0003	0.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB 88853

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I4344

WorkList ID : 100927

Date : 7/21/2017 8:35:23 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4344-01-D	TSS	Cool 4 deg C	SKAN06	H21	SPDES-1	07/20/2017	SM2540 D

Date/Time 07/21/17 8:35 AM
 Received by: JP
 Relinquished by: JP

Date/Time 07/21/17 10:00 AM
 Received by: CP
 Relinquished by: JP