

7470A


LB88865

INSTRUMENT ID: CYJ

Linear

 μ Abs.:

3272E



A= 0.0000e+000

B= 3.0683e-004

C= 3.8513e-002

Rho= 0.9998013

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-32	0.000	-32				
0.2	0.200	0.228	0.028	618	0.0 %	618				
2.5	2.500	2.537	0.037	8144	0.0 %	8144				
5.0	5.000	4.850	-0.150	15681	0.0 %	15681				
7.5	7.500	7.476	-0.024	24241	0.0 %	24241				
10.0	10.000	10.080	0.080	32726	0.0 %	32726				

LB88865 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jul 2017 11:26:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-32	-	0.0000	7470A	PPB	21 Jul 2017 11:28:55	S	Std
0.2 - 1	50.2	618	-	0.2000	7470A	PPB	21 Jul 2017 11:31:03	S	Std
2.5 - 1	52.5	8144	-	2.5000	7470A	PPB	21 Jul 2017 11:33:11	S	Std
5.0 - 1	55	15681	-	5.0000	7470A	PPB	21 Jul 2017 11:35:20	S	Std
7.5 - 1	57.5	24241	-	7.5000	7470A	PPB	21 Jul 2017 11:37:27	S	Std
10.0 - 1	510	32726	-	10.0000	7470A	PPB	21 Jul 2017 11:39:35	S	Std
ICV70 - 1	ICV70	12394	3.8413	-7470A	PPB	21 Jul 2017 11:42:13	U	SMPL	
ICB70 - 1	ICB70	-126	-0.0001	-7470A	PPB	21 Jul 2017 11:44:19	U	SMPL	
CCV65 - 1	CCV65	15954	4.9336	-7470A	PPB	21 Jul 2017 11:46:26	U	SMPL	
CCB65 - 1	CCB65	-127	-0.0005	-7470A	PPB	21 Jul 2017 11:48:33	U	SMPL	
CRA - 1	CRA	593	0.2205	-7470A	PPB	21 Jul 2017 11:50:40	U	SMPL	
HighStd - 1	HighStd	32563	10.0297	-7470A	PPB	21 Jul 2017 11:52:46	U	SMPL	
ChkStd - 1	ChkStd	22514	6.9464	-7470A	PPB	21 Jul 2017 11:54:53	U	SMPL	
PB100823BL - 1	PBW	-227	-0.0311	-7470A	PPB	21 Jul 2017 11:57:01	U	SMPL	
PB100823BS - 1	LCSW	13128	4.0665	-7470A	PPB	21 Jul 2017 11:59:10	U	SMPL	
PB100823TB - 1	PB100771TB	-145	-0.0060	-7470A	PPB	21 Jul 2017 12:01:18	U	SMPL	
I4044-01 - 1	SB-03-COMP(0-4.0)	-15	0.0339	-7470A	PPB	21 Jul 2017 12:03:26	U	SMPL	
I4280-03 - 1	CL-01-071817-A	67	0.0591	-7470A	PPB	21 Jul 2017 12:05:34	U	SMPL	
I4280-06 - 1	CL-01-071817-B	-9	0.0358	-7470A	PPB	21 Jul 2017 12:07:42	U	SMPL	
I4280-09 - 1	CL-01-071817-C	-18	0.0330	-7470A	PPB	21 Jul 2017 12:09:49	U	SMPL	
CCV66 - 1	CCV66	15974	4.9398	-7470A	PPB	21 Jul 2017 12:11:56	U	SMPL	
CCB66 - 1	CCB66	-125	0.0002	-7470A	PPB	21 Jul 2017 12:14:03	U	SMPL	
I4280-12 - 1	CL-01-071817-D	-33	0.0284	-7470A	PPB	21 Jul 2017 12:16:11	U	SMPL	
I4308-03 - 1	NB-08-071917	-46	0.0244	-7470A	PPB	21 Jul 2017 12:18:17	U	SMPL	
I4312-03 - 1	VNJ-209	-12	0.0348	-7470A	PPB	21 Jul 2017 12:20:24	U	SMPL	
I4312-07 - 1	72-11987	-18	0.0330	-7470A	PPB	21 Jul 2017 12:22:31	U	SMPL	
I4312-11 - 1	72-11978	-12	0.0348	-7470A	PPB	21 Jul 2017 12:24:39	U	SMPL	
I4312-15 - 1	VNJ-237	-4	0.0373	-7470A	PPB	21 Jul 2017 12:26:47	U	SMPL	
I4312-19 - 1	156	-23	0.0315	-7470A	PPB	21 Jul 2017 12:28:55	U	SMPL	
I4313-04 - 1	B-COMB-1	3	0.0394	-7470A	PPB	21 Jul 2017 12:31:03	U	SMPL	
I4316-03 - 1	OR-02-071917	-27	0.0302	-7470A	PPB	21 Jul 2017 12:33:12	U	SMPL	
I4317-03 - 1	SP-1-TP-1-2-3-COMP	-18	0.0330	-7470A	PPB	21 Jul 2017 12:35:19	U	SMPL	
CCV67 - 1	CCV67	15857	4.9039	-7470A	PPB	21 Jul 2017 12:37:25	U	SMPL	
CCB67 - 1	CCB67	-108	0.0054	-7470A	PPB	21 Jul 2017 12:39:32	U	SMPL	
I4317-03DUP - 1	SP-1-TP-1-2-3-COMPDUP	-3	0.0376	-7470A	PPB	21 Jul 2017 12:41:40	U	SMPL	
I4317-03MS - 1	SP-1-TP-1-2-3-COMPMS	11899	3.6894	-7470A	PPB	21 Jul 2017 12:43:46	U	SMPL	
I4317-03MSD - 1	SP-1-TP-1-2-3-COMPMSD	12346	3.8266	-7470A	PPB	21 Jul 2017 12:45:53	U	SMPL	
I4317-03LX5 - 1		-97	0.0088	-7470A	PPB	21 Jul 2017 12:48:01	U	SMPL	
I4317-03A - 1		12334	3.8229	-7470A	PPB	21 Jul 2017 12:50:08	U	SMPL	
CCV68 - 1	CCV68	15716	4.8606	-7470A	PPB	21 Jul 2017 12:52:16	U	SMPL	
CCB68 - 1	CCB68	-124	0.0005	-7470A	PPB	21 Jul 2017 12:54:24	U	SMPL	