



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88875

Temp.: 21.6 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/21/2017	12:15
2	CAL2	1	Water			7.00	07/21/2017	12:16
3	CAL3	1	Water			10.03	07/21/2017	12:17
4	ICV	1	Water			7.01	07/21/2017	12:19
5	CCV1	1	Water			2.02	07/21/2017	12:20
6	I4362-02	1	Water			6.50	07/21/2017	12:21
7	I4362-02DUP	1	Water			6.51	07/21/2017	12:22
8	I4369-08	1	Water			7.16	07/21/2017	12:25
9	CCV2	1	Water			12.01	07/21/2017	12:26

WORKLIST(Hardcopy Internal Chain)

LB 88875

WorkList Name : CORROSIONITY i4369

WorkList ID : 100953

Date : 7/21/2017 12:00:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2017	Water	I4362-02	-F Corrosivity	Cool 4 deg C	PSEG01	H41	FRAC-TANK	07/20/2017	9040C
	Water	I4369-08	-D Corrosivity	Cool 4 deg C	AECO05		WC-1-20170720	07/20/2017	9040C

Date/Time 07/21/17 12:00 PM
 Received by: 30
 Relinquished by:

Date/Time 07/21/17 2:00 PM
 Received by: CF
 Relinquished by: 70