


7471A

LB88881
INSTRUMENT ID: CV2

Linear

 μ Abs.:

41297



A= 0.0000e+000

B= 2.4557e-004

C= 2.2512e-002

Rho= 0.9993930

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	89	0.000	89				
0.2	0.200	0.220	0.020	805	0.0 %	805				
2.5	2.500	2.481	-0.019	10010	0.0 %	10010				
5.0	5.000	5.050	0.050	20473	0.0 %	20473				
7.5	7.500	7.241	-0.259	29395	0.0 %	29395	0			
10.0	10.000	10.164	0.164	41297	0.0 %	41297				

LB88881 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 21 Jul 2017 15:51:38

Sample ID	Extended ID	μ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	89	-	0.00007471A	PPB		21 Jul 2017 15:57:17	S	Std
0.2 - 1	50.2	805	-	0.20007471A	PPB		21 Jul 2017 15:59:34	S	Std
2.5 - 1	52.5	10010	-	2.50007471A	PPB		21 Jul 2017 16:01:52	S	Std
5.0 - 1	59	20473	-	5.00007471A	PPB		21 Jul 2017 16:06:55	S	Std
7.5 - 1	57.5	29395	-	7.50007471A	PPB		21 Jul 2017 16:09:12	S	Std
10.0 - 1	510	41297	-	10.00007471A	PPB		21 Jul 2017 16:14:02	S	Std
ICV71 - 1	ICV71	15259	3.7696	-7471A	PPB		21 Jul 2017 16:16:42	U	SMPL
ICB71 - 1	ICB71	-119	-0.0067	-7471A	PPB		21 Jul 2017 16:18:57	U	SMPL
CCV71 - 1	CCV71	20317	5.0117	-7471A	PPB		21 Jul 2017 16:21:13	U	SMPL
CCB71 - 1	CCB71	-137	-0.0111	-7471A	PPB		21 Jul 2017 16:23:29	U	SMPL
CRA - 1	CRA	774	0.2126	-7471A	PPB		21 Jul 2017 16:25:45	U	SMPL
HighStd - 1	HighStd	41871	10.3047	-7471A	PPB		21 Jul 2017 16:28:01	U	SMPL
ChkStd - 1	ChkStd	27920	6.8788	-7471A	PPB		21 Jul 2017 16:30:16	U	SMPL
PB100820BL - 1	PBS	-119	-0.0067	-7471A	PPB		21 Jul 2017 16:32:34	U	SMPL
PB100820BS - 1	LCSS	14887	3.6783	-7471A	PPB		21 Jul 2017 16:34:52	U	SMPL
I4244-01 - 1	PENNSYLVANIA-ANALYTICAL	12380	3.0626	-7471A	PPB		21 Jul 2017 16:37:09	U	SMPL
I4302-02 - 1	SB-2D1011	5433	1.3567	-7471A	PPB		21 Jul 2017 16:39:26	U	SMPL
I4302-02DUP - 1	SB-2D1011DUP	5076	1.2690	-7471A	PPB		21 Jul 2017 16:41:44	U	SMPL
I4302-02MS - 1	SB-2D1011MS	21671	5.3442	-7471A	PPB		21 Jul 2017 16:44:03	U	SMPL
I4302-02MSD - 1	SB-2D1011MSD	20926	5.1613	-7471A	PPB		21 Jul 2017 16:51:09	U	SMPL
CCV72 - 1	CCV72	19980	4.9290	-7471A	PPB		21 Jul 2017 16:53:25	U	SMPL
CCB72 - 1	CCB72	-144	-0.0128	-7471A	PPB		21 Jul 2017 16:55:42	U	SMPL
I4302-03 - 1	SB-4D12	16421	4.0550	-7471A	PPB		21 Jul 2017 16:57:58	U	SMPL
I4302-04 - 1	SB-7D1213	301	0.0964	-7471A	PPB		21 Jul 2017 17:00:14	U	SMPL
I4302-06 - 1	SB-11D1011	566	0.1615	-7471A	PPB		21 Jul 2017 17:02:30	U	SMPL
I4310-01 - 1	B3-10GRIT-COMP	4	0.0235	-7471A	PPB		21 Jul 2017 17:04:46	U	SMPL
I4347-02 - 1	B-2(6-10)COMP	44865	11.0399	-7471A	PPB		21 Jul 2017 17:07:02	U	SMPL
I4352-01 - 1	FK-SF-03-001-A	1794	0.4631	-7471A	PPB		21 Jul 2017 17:09:19	U	SMPL
I4352-03 - 1	FK-SF-03-001-B	1271	0.3346	-7471A	PPB		21 Jul 2017 17:11:38	U	SMPL
I4352-05 - 1	FK-SF-03-001-C	1457	0.3803	-7471A	PPB		21 Jul 2017 17:13:55	U	SMPL
I4358-01 - 1	TR-05-072017	1310	0.3442	-7471A	PPB		21 Jul 2017 17:16:13	U	SMPL
I4361-01 - 1	RAMP-D-1(0-2)	1186	0.3138	-7471A	PPB		21 Jul 2017 17:18:29	U	SMPL
CCV73 - 1	CCV73	20134	4.9668	-7471A	PPB		21 Jul 2017 17:20:46	U	SMPL
CCB73 - 1	CCB73	-96	-0.0011	-7471A	PPB		21 Jul 2017 17:23:02	U	SMPL
I4361-03 - 1	RAMP-D-2(0-2)	744	0.2052	-7471A	PPB		21 Jul 2017 17:25:19	U	SMPL
I4361-05 - 1	RAMP-D-3(0-2)	758	0.2087	-7471A	PPB		21 Jul 2017 17:27:35	U	SMPL
I4361-07 - 1	RAMP-D-4(0-2)	467	0.1372	-7471A	PPB		21 Jul 2017 17:29:52	U	SMPL
I4361-09 - 1	RAMP-D-5(0-2)	657	0.1839	-7471A	PPB		21 Jul 2017 17:32:08	U	SMPL
I4370-05 - 1	E2-N6-BASE	61835	15.2072	-7471A	PPB		21 Jul 2017 17:34:25	U	SMPL
I4370-06 - 1	E2-N5-NWSW	204416	50.2206	-7471A	PPB		21 Jul 2017 17:36:42	U	SMPL
I4370-07 - 1	E2-N5-SWSW	56857	13.9848	-7471A	PPB		21 Jul 2017 17:39:02	U	SMPL
PB100822BL - 1	PBS	-459	-0.0902	-7471A	PPB		21 Jul 2017 17:41:32	U	SMPL
PB100822BS - 1	LCSS	14420	3.5636	-7471A	PPB		21 Jul 2017 17:43:52	U	SMPL
I4359-01 - 1	CEMETERY-1(0-2)	1963	0.5046	-7471A	PPB		21 Jul 2017 17:46:09	U	SMPL
CCV74 - 1	CCV74	19952	4.9221	-7471A	PPB		21 Jul 2017 17:48:25	U	SMPL
CCB74 - 1	CCB74	-86	0.0014	-7471A	PPB		21 Jul 2017 17:50:42	U	SMPL
I4359-03 - 1	CEMETERY-2(0-2)	1434	0.3747	-7471A	PPB		21 Jul 2017 17:52:59	U	SMPL
I4359-05 - 1	CEMETERY-3(0-2)	919	0.2482	-7471A	PPB		21 Jul 2017 17:55:15	U	SMPL
I4359-07 - 1	CEMETERY-4(0-2)	712	0.1974	-7471A	PPB		21 Jul 2017 17:57:32	U	SMPL
I4359-09 - 1	CEMETERY-5(0-2)	4877	1.2201	-7471A	PPB		21 Jul 2017 17:59:48	U	SMPL
I4359-11 - 1	CEMETERY-6(0-2)	5534	1.3815	-7471A	PPB		21 Jul 2017 18:02:05	U	SMPL
I4359-13 - 1	CEMETERY-7(0-2)	1081	0.2880	-7471A	PPB		21 Jul 2017 18:04:22	U	SMPL
I4359-15 - 1	CEMETERY-8(0-2)	1392	0.3643	-7471A	PPB		21 Jul 2017 18:06:39	U	SMPL
I4359-17 - 1	CEMETERY-9(0-2)	9538	2.3647	-7471A	PPB		21 Jul 2017 18:08:57	U	SMPL
I4359-19 - 1	CEMETERY-10(0-2)	1634	0.4238	-7471A	PPB		21 Jul 2017 18:11:14	U	SMPL
I4360-01 - 1	CEMETERY-11(0-2)	7254	1.8039	-7471A	PPB		21 Jul 2017 18:13:31	U	SMPL
CCV75 - 1	CCV75	20241	4.9931	-7471A	PPB		21 Jul 2017 18:15:47	U	SMPL
CCB75 - 1	CCB75	-92	-0.0001	-7471A	PPB		21 Jul 2017 18:18:03	U	SMPL
I4360-03 - 1	CEMETERY-12(0-2)	2098	0.5377	-7471A	PPB		21 Jul 2017 18:20:20	U	SMPL
I4360-05 - 1	CEMETERY-13(0-2)	4029	1.0119	-7471A	PPB		21 Jul 2017 18:22:37	U	SMPL
I4360-07 - 1	CEMETERY-14(0-2)	2281	0.5827	-7471A	PPB		21 Jul 2017 18:24:53	U	SMPL
I4360-09 - 1	CEMETERY-15(0-2)	1643	0.4260	-7471A	PPB		21 Jul 2017 18:27:10	U	SMPL
I4360-11 - 1	295-N-1(0-2)	492	0.1433	-7471A	PPB		21 Jul 2017 18:29:26	U	SMPL
I4360-13 - 1	295-N-2(0-2)	398	0.1202	-7471A	PPB		21 Jul 2017 18:31:43	U	SMPL
I4360-15 - 1	295-N-3(0-2)	1535	0.3995	-7471A	PPB		21 Jul 2017 18:34:00	U	SMPL
I4360-17 - 1	295-N-4(0-2)	1147	0.3042	-7471A	PPB		21 Jul 2017 18:36:18	U	SMPL
I4360-19 - 1	295-N-5(0-2)	812	0.2219	-7471A	PPB		21 Jul 2017 18:38:36	U	SMPL
I4360-19DUP - 1	295-N-5(0-2)DUP	711	0.1971	-7471A	PPB		21 Jul 2017 18:40:52	U	SMPL
CCV76 - 1	CCV76	20102	4.9589	-7471A	PPB		21 Jul 2017 18:43:09	U	SMPL
CCB76 - 1	CCB76	-96	-0.0011	-7471A	PPB		21 Jul 2017 18:45:26	U	SMPL
I4360-19MS - 1	295-N-5(0-2)MS	15461	3.8192	-7471A	PPB		21 Jul 2017 18:47:43	U	SMPL
I4360-19MSD - 1	295-N-5(0-2)MSD	14781	3.6523	-7471A	PPB		21 Jul 2017 18:49:59	U	SMPL
I4302-02LX5 - 1		1020	0.2730	-7471A	PPB		21 Jul 2017 18:52:16	U	SMPL
I4302-02A - 1		20701	5.1060	-7471A	PPB		21 Jul 2017 18:54:32	U	SMPL
I4360-19LX5 - 1		90	0.0446	-7471A	PPB		21 Jul 2017 18:56:49	U	SMPL
I4360-19A - 1		16340	4.0351	-7471A	PPB		21 Jul 2017 18:59:06	U	SMPL
I4347-02DLX2 - 1		22014	5.4285	-7471A	PPB		21 Jul 2017 19:01:24	U	SMPL
I4370-05DLX5 - 1		12094	2.9924	-7471A	PPB		21 Jul 2017 19:03:41	U	SMPL
I4370-06DLX10 - 1		20733	5.1139	-7471A	PPB		21 Jul 2017 19:05:59	U	SMPL
I4370-07DLX2 - 1		28086	6.9195	-7471A	PPB		21 Jul 2017 19:08:16	U	SMPL
CCV77 - 1	CCV77	19961	4.9243	-7471A	PPB		21 Jul 2017 19:10:34	U	SMPL
CCB77 - 1	CCB77	-124	-0.0079	-7471A	PPB		21 Jul 2017 19:12:52	U	SMPL