



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88882

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/21/2017	14:00
2	CAL2	1	Water			6.99	07/21/2017	14:01
3	CAL3	1	Water			10.02	07/21/2017	14:02
4	ICV	1	Water			7.00	07/21/2017	14:05
5	CCV1	1	Water			2.01	07/21/2017	14:06
6	I4378-02	1	Solid	20.02	20	5.88	07/21/2017	14:07
7	I4378-02DUP	1	Solid	20.03	20	5.89	07/21/2017	14:08
8	I4378-04	1	Solid	20.01	20	6.91	07/21/2017	14:10
9	I4378-06	1	Solid	20.03	20	6.53	07/21/2017	14:11
10	CCV2	1	Water			12.01	07/21/2017	14:13

LB88882

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4378 WorkList ID : 100955 Date : 7/21/2017 12:03:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/27/2017	Solid	I4378-02	Corrosivity	Cool 4 deg C	PSEG02	H42	DRUM-1	07/20/2017	9045C
07/27/2017	Solid	I4378-04	Corrosivity	Cool 4 deg C	PSEG02	H42	DRUM-2	07/20/2017	9045C
07/27/2017	Solid	I4378-06	Corrosivity	Cool 4 deg C	PSEG02	H42	DRUM-3	07/20/2017	9045C

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: