



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88900

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/21/2017	16:00
2	CAL2	1	Water			7.01	07/21/2017	16:01
3	CAL3	1	Water			10.02	07/21/2017	16:03
4	ICV	1	Water			6.99	07/21/2017	16:05
5	CCV1	1	Water			2.02	07/21/2017	16:06
6	I4380-03	1	Solid	20.02	20	8.24	07/21/2017	16:07
7	I4391-03	1	Solid	20.03	20	5.82	07/21/2017	16:09
8	I4391-06	1	Solid	20.04	20	6.80	07/21/2017	16:10
9	I4391-06DUP	1	Solid	20.02	20	6.81	07/21/2017	16:12
10	CCV2	1	Water			12.01	07/21/2017	16:13

2B88900

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4391 WorkList ID : 100963 Date : 7/21/2017 2:17:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4380-03	Corrosivity	Cool 4 deg C	WSP001	H32	B3(0-10)COMP	07/21/2017	9045C
07/26/2017	Solid	I4391-03	Corrosivity	Cool 4 deg C	PSEG05	H42	BU-02-072117-A	07/21/2017	9045C
07/26/2017	Solid	I4391-06	Corrosivity	Cool 4 deg C	PSEG05	H42	BU-02-072117-B	07/21/2017	9045C

Date/Time 07/21/17 01:20 PM
Received by: JA
Relinquished by: CS

Date/Time 07/21/17 5:10 PM
Received by: CP
Relinquished by: JA