



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88903

Temp.: 23.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/21/2017	19:00
2	CAL2	1	Water			7.00	07/21/2017	19:01
3	CAL3	1	Water			10.03	07/21/2017	19:02
4	ICV	1	Water			6.99	07/21/2017	19:05
5	CCV1	1	Water			2.02	07/21/2017	19:06
6	I4402-08	1	Water			6.83	07/21/2017	19:08
7	I4402-08DUP	1	Water			6.84	07/21/2017	19:10
8	CCV2	1	Water			12.01	07/21/2017	19:11

WORKLIST(Hardcopy Internal Chain)

LB 88903

WorkList Name : CORROSIVITY I4402 WorkList ID : 100973 Date : 7/21/2017 6:39:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/21/2017	Water	I4402-08~1	Corrosivity	Cool 4 deg C	PSEG01	H53	FRAC-TANK-260783	07/21/2017	9040C

Date/Time 07/21/17 6:40 PM
 Received by: JA
 Relinquished by: CB

Date/Time 07/21/17 7:20 PM
 Received by: CP
 Relinquished by: JP