



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 07/25/2017DATE & TIME IN: 07/24/2017 17:10OVEN TEMP: 98 °CRun Number: LB88925DATE & TIME OUT: 07/25/2017 12:50OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-02

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB88925BL	LB88925BL	89.6037	89.6037	100	89.6037	89.6037	89.6037	0	0
2	I4363-01	B-COMB-3	83.5605	83.5605	100	83.6102	83.6102	83.6102	0.0497	497
3	I4363-01DUP	B-COMB-3DUP	81.4331	81.4331	100	81.4833	81.4833	81.4833	0.0502	502
4	I4363-02	B-COMB-4	79.5235	79.5235	100	79.6013	79.6013	79.6013	0.0778	778

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

288925

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I4363

WorkList ID : 101033

Date : 7/24/2017 8:30:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4363-01	TS	Cool 4 deg C	DIST01	H32	B-COMB-3	07/20/2017	SM2540 B
	Water	I4363-02	TS	Cool 4 deg C	DIST01	H32	B-COMB-4	07/20/2017	SM2540 B

Date/Time 07/24/17 11:00 AM
 Received by: JR
 Relinquished by: CP

Date/Time 07/24/17 5:30 PM
 Received by: CP
 Relinquished by: JD