



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88929

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	07/19/2017	16:00
2	CAL2	1	Water			7.01	07/19/2017	16:01
3	CAL3	1	Water			10.03	07/19/2017	16:03
4	ICV	1	Water			7.00	07/19/2017	16:05
5	CCV1	1	Water			2.02	07/19/2017	16:06
6	I4280-03	1	Solid	20.03	20	8.84	07/19/2017	16:09
7	I4280-03DUP	1	Solid	20.04	20	8.85	07/19/2017	16:10
8	I4280-06	1	Solid	20.02	20	7.36	07/19/2017	16:12
9	I4280-09	1	Solid	20.01	20	8.02	07/19/2017	16:13
10	I4280-12	1	Solid	20.04	20	8.48	07/19/2017	16:15
11	I4308-03	1	Solid	20.05	20	8.87	07/19/2017	16:16
12	I4312-03	1	Solid	20.02	20	8.31	07/19/2017	16:17
13	I4312-07	1	Solid	20.03	20	7.82	07/19/2017	16:19
14	I4312-11	1	Solid	20.02	20	7.93	07/19/2017	16:22
15	I4312-15	1	Solid	20.01	20	7.76	07/19/2017	16:25
16	CCV2	1	Water			12.01	07/19/2017	16:26
17	I4312-19	1	Solid	20.02	20	7.90	07/19/2017	16:27
18	I4313-04	1	Solid	20.03	20	8.75	07/19/2017	16:30
19	I4316-03	1	Solid	20.01	20	7.93	07/19/2017	16:32
20	I4317-03	1	Solid	20.03	20	6.90	07/19/2017	16:34

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I4319-02	1	Solid	20.04	20	6.91	07/19/2017	16:35
22	I4319-02DUP	1	Solid	20.04	20	6.92	07/19/2017	16:36
23	CCV3	1	Water			12.01	07/19/2017	16:37