


245.1

LB 88937
INSTRUMENT ID: CV2

Linear

 μ Abs.:

30407



A= 0.0000e+000

B= 3.3170e-004

C= -1.9602e-002

Rho= 0.9998966

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	153	0.000	153				
0.2	0.200	0.233	0.033	762	0.0 %	762				
2.5	2.500	2.491	-0.009	7569	0.0 %	7569				
5.0	5.000	4.901	-0.099	14834	0.0 %	14834				
7.5	7.500	7.477	-0.023	22602	0.0 %	22602				
10.0	10.000	10.066	0.066	30407	0.0 %	30407				

LB88937 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Jul 2017 14:27:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	153	-	0.0000	245.1	PPB	24 Jul 2017 14:29:30	S	Std
0.2 - 1	50.2	762	-	0.2000	245.1	PPB	24 Jul 2017 14:31:39	S	Std
2.5 - 1	52.5	7569	-	2.5000	245.1	PPB	24 Jul 2017 14:33:47	S	Std
5.0 - 1	55	14834	-	5.0000	245.1	PPB	24 Jul 2017 14:35:55	S	Std
7.5 - 1	57.5	22602	-	7.5000	245.1	PPB	24 Jul 2017 14:38:03	S	Std
10.0 - 1	510	30407	-	10.0000	245.1	PPB	24 Jul 2017 14:42:49	S	Std
ICV73 - 1	ICV73	11687	3.8570	-245.1		PPB	24 Jul 2017 14:45:20	U	SMPL
ICB73 - 1	ICB73	14	-0.0150	-245.1		PPB	24 Jul 2017 14:47:27	U	SMPL
CCV84 - 1	CCV84	14880	4.9161	-245.1		PPB	24 Jul 2017 14:49:34	U	SMPL
CCB84 - 1	CCB84	2	-0.0189	-245.1		PPB	24 Jul 2017 14:51:40	U	SMPL
CRA - 1	CRA	746	0.2278	-245.1		PPB	24 Jul 2017 14:53:48	U	SMPL
HighStd - 1	HighStd	30027	9.9403	-245.1		PPB	24 Jul 2017 14:55:54	U	SMPL
ChkStd - 1	ChkStd	20480	6.7736	-245.1		PPB	24 Jul 2017 14:58:01	U	SMPL
PB100855BL - 1	PBW	-52	-0.0369	-245.1		PPB	24 Jul 2017 15:00:11	U	SMPL
PB100855BS - 1	LCSW	12223	4.0348	-245.1		PPB	24 Jul 2017 15:02:20	U	SMPL
I4363-01 - 1	B-COMB-3	82	0.0076	-245.1		PPB	24 Jul 2017 15:04:28	U	SMPL
I4363-01DUP - 1	B-COMB-3DUP	162	0.0341	-245.1		PPB	24 Jul 2017 15:06:36	U	SMPL
I4363-01MS - 1	B-COMB-3MS	11228	3.7047	-245.1		PPB	24 Jul 2017 15:08:44	U	SMPL
I4363-01MSD - 1	B-COMB-3MSD	11152	3.6795	-245.1		PPB	24 Jul 2017 15:13:45	U	SMPL
I4363-02 - 1	B-COMB-4	319	0.0862	-245.1		PPB	24 Jul 2017 15:15:51	U	SMPL
CCV85 - 1	CCV85	14829	4.8992	-245.1		PPB	24 Jul 2017 15:17:59	U	SMPL
CCB85 - 1	CCB85	0	-0.0196	-245.1		PPB	24 Jul 2017 15:20:05	U	SMPL
I4363-01LX5 - 1		106	0.0156	-245.1		PPB	24 Jul 2017 15:22:13	U	SMPL
I4363-01A - 1		12485	4.1217	-245.1		PPB	24 Jul 2017 15:24:19	U	SMPL
CCV86 - 1	CCV86	14727	4.8653	-245.1		PPB	24 Jul 2017 15:26:26	U	SMPL
CCB86 - 1	CCB86	16	-0.0143	-245.1		PPB	24 Jul 2017 15:28:33	U	SMPL