


7471A

LB 88944
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40667



A= 0.0000e+000

B= 2.4987e-004

C= 1.8817e-002

Rho= 0.9992862

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	63	0.000	63				
0.2	0.200	0.230	0.030	845	0.0 %	845				
2.5	2.500	2.498	-0.002	9920	0.0 %	9920				
5.0	5.000	5.041	0.041	20098	0.0 %	20098				
7.5	7.500	7.217	-0.283	28806	0.0 %	28806	0			
10.0	10.000	10.180	0.180	40667	0.0 %	40667				

LB88944 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 24 Jul 2017 15:55:31					
Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.0000	7471A	PPB	24 Jul 2017 15:58:00	S	Std
0.2 - 1	50.2	845	-	0.2000	7471A	PPB	24 Jul 2017 16:00:18	S	Std
2.5 - 1	52.5	9920	-	2.5000	7471A	PPB	24 Jul 2017 16:02:36	S	Std
5.0 - 1	55	20098	-	5.0000	7471A	PPB	24 Jul 2017 16:04:54	S	Std
7.5 - 1	57.5	28806	-	7.5000	7471A	PPB	24 Jul 2017 16:07:11	S	Std
10.0 - 1	510	40667	-	10.0000	7471A	PPB	24 Jul 2017 16:12:03	S	Std
ICV74 - 1	ICV74	15582	3.9124	-	7471A	PPB	24 Jul 2017 16:14:47	U	SMPL
ICB74 - 1	ICB74	-101	-0.0064	-	7471A	PPB	24 Jul 2017 16:17:02	U	SMPL
CCV87 - 1	CCV87	20066	5.0328	-	7471A	PPB	24 Jul 2017 16:19:18	U	SMPL
CCB87 - 1	CCB87	-101	-0.0064	-	7471A	PPB	24 Jul 2017 16:21:34	U	SMPL
CRA - 1	CRA	823	0.2245	-	7471A	PPB	24 Jul 2017 16:23:50	U	SMPL
HighStd - 1	HighStd	40760	10.2037	-	7471A	PPB	24 Jul 2017 16:26:06	U	SMPL
ChkStd - 1	ChkStd	27745	6.9516	-	7471A	PPB	24 Jul 2017 16:28:22	U	SMPL
PB100858BL - 1	PBS	-191	-0.0289	-	7471A	PPB	24 Jul 2017 16:30:41	U	SMPL
PB100858BS - 1	LCSS	14524	3.6480	-	7471A	PPB	24 Jul 2017 16:32:59	U	SMPL
I4380-02 - 1	B3(0-10)COMP	4310	1.0958	-	7471A	PPB	24 Jul 2017 16:35:16	U	SMPL
I4380-02DUP - 1	B3(0-10)COMP DUP	4079	1.0381	-	7471A	PPB	24 Jul 2017 16:37:33	U	SMPL
I4380-02MS - 1	B3(0-10)COMPMS	18254	4.5800	-	7471A	PPB	24 Jul 2017 16:39:51	U	SMPL
I4380-02MSD - 1	B3(0-10)COMPMSD	18909	4.7437	-	7471A	PPB	24 Jul 2017 16:42:09	U	SMPL
I4391-01 - 1	BU-02-072117-A	715	0.1975	-	7471A	PPB	24 Jul 2017 16:44:25	U	SMPL
CCV88 - 1	CCV88	20149	5.0535	-	7471A	PPB	24 Jul 2017 16:46:41	U	SMPL
CCB88 - 1	CCB88	-131	-0.0139	-	7471A	PPB	24 Jul 2017 16:48:57	U	SMPL
I4391-04 - 1	BU-02-072117-B	1368	0.3606	-	7471A	PPB	24 Jul 2017 16:51:14	U	SMPL
I4398-01 - 1	SEDIMENT-1	8826	2.2242	-	7471A	PPB	24 Jul 2017 16:53:30	U	SMPL
I4399-02 - 1	E2-N7-BASE	402	0.1193	-	7471A	PPB	24 Jul 2017 16:55:46	U	SMPL
I4399-03 - 1	E2-HS4-WSW	2914	0.7469	-	7471A	PPB	24 Jul 2017 16:58:02	U	SMPL
I4399-04 - 1	E2-HS4-SSW	1878	0.4881	-	7471A	PPB	24 Jul 2017 17:00:19	U	SMPL
I4399-05 - 1	E2-HS4-ESW	21859	5.4808	-	7471A	PPB	24 Jul 2017 17:02:36	U	SMPL
I4399-06 - 1	E2-HS4-NSW	11478	2.8869	-	7471A	PPB	24 Jul 2017 17:04:54	U	SMPL
I4399-07 - 1	E2-HS4-BASE	6260	1.5830	-	7471A	PPB	24 Jul 2017 17:07:12	U	SMPL
I4380-02LX5 - 1		831	0.2265	-	7471A	PPB	24 Jul 2017 17:09:30	U	SMPL
I4380-02A - 1		20132	5.0493	-	7471A	PPB	24 Jul 2017 17:11:46	U	SMPL
CCV89 - 1	CCV89	20072	5.0343	-	7471A	PPB	24 Jul 2017 17:14:02	U	SMPL
CCB89 - 1	CCB89	-92	-0.0042	-	7471A	PPB	24 Jul 2017 17:16:18	U	SMPL