

7470A


LB88973

INSTRUMENT ID: CY1

Linear

 μ Abs.:

33114



A= 0.0000e+000

B= 3.0015e-004

C= -3.1544e-003

Rho= 0.9998857

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	7	0.000	7				
0.2	0.200	0.215	0.015	727	0.0 %	727				
2.5	2.500	2.422	-0.078	8080	0.0 %	8080				
5.0	5.000	5.059	0.059	16864	0.0 %	16864				
7.5	7.500	7.570	0.070	25230	0.0 %	25230				
10.0	10.000	9.936	-0.064	33114	0.0 %	33114				

LB88973 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jul 2017 10:56:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	25 Jul 2017 11:02:03	S	Std
0.2 - 1	50.2	727	-	0.2000	7470A	PPB	25 Jul 2017 11:04:11	S	Std
2.5 - 1	52.5	8080	-	2.5000	7470A	PPB	25 Jul 2017 11:06:19	S	Std
5.0 - 1	55	16864	-	5.0000	7470A	PPB	25 Jul 2017 11:08:28	S	Std
7.5 - 1	57.5	25230	-	7.5000	7470A	PPB	25 Jul 2017 11:10:36	S	Std
10.0 - 1	510	33114	-	10.0000	7470A	PPB	25 Jul 2017 11:12:44	S	Std
ICV75 - 1	ICV75	12163	3.6475	-7470A	PPB	25 Jul 2017 11:15:31	U	SMPL	
ICB75 - 1	ICB75	-126	-0.0410	-7470A	PPB	25 Jul 2017 11:17:38	U	SMPL	
CCV90 - 1	CCV90	16710	5.0123	-7470A	PPB	25 Jul 2017 11:19:44	U	SMPL	
CCB90 - 1	CCB90	-105	-0.0347	-7470A	PPB	25 Jul 2017 11:21:51	U	SMPL	
CRA - 1	CRA	690	0.2039	-7470A	PPB	25 Jul 2017 11:23:58	U	SMPL	
HighStd - 1	HighStd	33110	9.9347	-7470A	PPB	25 Jul 2017 11:26:05	U	SMPL	
ChkStd - 1	ChkStd	22885	6.8657	-7470A	PPB	25 Jul 2017 11:28:12	U	SMPL	
PB100915BL - 1	PBW	-184	-0.0584	-7470A	PPB	25 Jul 2017 11:30:21	U	SMPL	
PB100915BS - 1	LCSW	12610	3.7817	-7470A	PPB	25 Jul 2017 11:32:30	U	SMPL	
PB100915TB - 1	PB100850TB	-99	-0.0329	-7470A	PPB	25 Jul 2017 11:34:38	U	SMPL	
I4362-02 - 1	FRAC-TANK	-15	-0.0077	-7470A	PPB	25 Jul 2017 11:36:47	U	SMPL	
I4369-08 - 1	WC-1-20170720	22	0.0034	-7470A	PPB	25 Jul 2017 11:38:55	U	SMPL	
I4380-03 - 1	B3(0-10)COMP	42	0.0095	-7470A	PPB	25 Jul 2017 11:41:03	U	SMPL	
I4380-03DUP - 1	B3(0-10)COMP DUP	20	0.0028	-7470A	PPB	25 Jul 2017 11:43:10	U	SMPL	
CCV91 - 1	CCV91	17114	5.1336	-7470A	PPB	25 Jul 2017 11:45:17	U	SMPL	
CCB91 - 1	CCB91	-121	-0.0395	-7470A	PPB	25 Jul 2017 11:47:24	U	SMPL	
I4380-03MS - 1	B3(0-10)COMPMS	12432	3.7283	-7470A	PPB	25 Jul 2017 11:49:31	U	SMPL	
I4380-03MSD - 1	B3(0-10)COMPMSD	12313	3.6926	-7470A	PPB	25 Jul 2017 11:51:38	U	SMPL	
I4391-03 - 1	BU-02-072117-A	-72	-0.0248	-7470A	PPB	25 Jul 2017 11:53:45	U	SMPL	
I4391-06 - 1	BU-02-072117-B	8	-0.0008	-7470A	PPB	25 Jul 2017 11:55:52	U	SMPL	
I4380-03LX5 - 1		34	0.0071	-7470A	PPB	25 Jul 2017 12:13:37	U	SMPL	
I4380-03A - 1		12701	3.8090	-7470A	PPB	25 Jul 2017 12:15:45	U	SMPL	
CCV92 - 1	CCV92	17118	5.1348	-7470A	PPB	25 Jul 2017 12:17:54	U	SMPL	
CCB92 - 1	CCB92	-142	-0.0458	-7470A	PPB	25 Jul 2017 12:20:01	U	SMPL	