



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB88984

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP56166
COD calibration std. 100 ppm	WP56164
COD calibration std. 10 ppm	WP56162
COD calibration std. 150 ppm	WP56165
COD calibration std. 50 ppm	WP56163
COD calibration std. 0 ppm	WP56161
COD Digestion Vials Low Level 0-150Mg/L	W2235
COD ICV-LCS std, 50ppm	WP56652
COD CCV std, 50ppm	WP56653

Temp In (C): <u>150</u>	Date In: <u>07/25/2017</u>	Time In: <u>09:10</u>
Temp Out (C): <u>150</u>	Date Out: <u>07/25/2017</u>	Time Out: <u>11:12</u>

Intercept: 0.2985

Slope: 1.0016

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/05/2017	16:00
2	CAL2	10	1	Water	10.000	07/05/2017	16:00
3	CAL3	50	1	Water	51.000	07/05/2017	16:01
4	CAL4	100	1	Water	101.000	07/05/2017	16:01
5	CAL5	150	1	Water	150.000	07/05/2017	16:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB88984

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.622	07/05/2017	16:02
2	ICB				1	Water	0.000	-0.298	07/05/2017	16:03
3	CCV1	50			1	Water	52.000	51.619	07/25/2017	12:01
4	CCB1				1	Water	0.000	-0.298	07/25/2017	12:01
5	LB88984BL				1	Water	0.000	-0.298	07/25/2017	12:02
6	LB88984BS	50			1	Water	51.000	50.621	07/25/2017	12:02
7	I4300-02				1	Water	0.000	-0.298	07/25/2017	12:03
8	I4300-02DUP				1	Water	0.000	-0.298	07/25/2017	12:03
9	I4300-02MS	50			1	Water	53.000	52.617	07/25/2017	12:04
10	I4300-02MSD	50			1	Water	52.000	51.619	07/25/2017	12:04
11	I4300-03				1	Water	3.000	2.697	07/25/2017	12:05
12	I4354-03				1	Water	38.000	37.641	07/25/2017	12:05
13	I4362-01				1	Water	120.000	119.510	07/25/2017	12:06
14	I4402-07				1	Water	16.000	15.676	07/25/2017	12:06
15	CCV2	50			1	Water	52.000	51.619	07/25/2017	12:07
16	CCB2				1	Water	0.000	-0.298	07/25/2017	12:07