


7471A

LB 82987
INSTRUMENT ID: CY4

Linear

 μ Abs.:

38228



A= 0.0000e+000

B= 2.6180e-004

C= -1.6053e-002

Rho= 0.9999483

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	135	0.000	135				
0.2	0.200	0.206	0.006	847	0.0 %	847				
2.5	2.500	2.495	-0.005	9591	0.0 %	9591				
5.0	5.000	4.931	-0.069	18897	0.0 %	18897				
7.5	7.500	7.557	0.057	28926	0.0 %	28926	0			
10.0	10.000	9.992	-0.008	38228	0.0 %	38228				

LB88987 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 25 Jul 2017 15:42:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	135	-	0.0000	7471A	PPB	25 Jul 2017 15:44:25	S	Std
0.2 - 1	50.2	847	-	0.2000	7471A	PPB	25 Jul 2017 15:46:43	S	Std
2.5 - 1	52.5	9591	-	2.5000	7471A	PPB	25 Jul 2017 15:49:00	S	Std
5.0 - 1	55	18897	-	5.0000	7471A	PPB	25 Jul 2017 15:51:18	S	Std
7.5 - 1	57.5	28926	-	7.5000	7471A	PPB	25 Jul 2017 15:53:35	S	Std
10.0 - 1	59.5	38228	-	10.0000	7471A	PPB	25 Jul 2017 16:01:06	S	Std
ICV77 - 1	ICV77	15481	4.0369	-	7471A	PPB	25 Jul 2017 16:03:46	U	SMPL
ICB77 - 1	ICB77	-115	-0.0462	-	7471A	PPB	25 Jul 2017 16:06:02	U	SMPL
CCV01 - 1	CCV01	19845	5.1794	-	7471A	PPB	25 Jul 2017 16:08:18	U	SMPL
CCB01 - 1	CCB01	-59	-0.0315	-	7471A	PPB	25 Jul 2017 16:10:33	U	SMPL
CRA - 1	CRA	818	0.1981	-	7471A	PPB	25 Jul 2017 16:12:49	U	SMPL
HighStd - 1	HighStd	38954	10.1822	-	7471A	PPB	25 Jul 2017 16:15:05	U	SMPL
ChkStd - 1	ChkStd	27379	7.1518	-	7471A	PPB	25 Jul 2017 16:17:21	U	SMPL
PB100928BL - 1	PBS	-22	-0.0218	-	7471A	PPB	25 Jul 2017 16:22:43	U	SMPL
PB100928BS - 1	LCSS	13881	3.6180	-	7471A	PPB	25 Jul 2017 16:24:59	U	SMPL
I4400-01 - 1	SB01A	12440	3.2408	-	7471A	PPB	25 Jul 2017 16:27:16	U	SMPL
I4400-02 - 1	SB01B	94	0.0086	-	7471A	PPB	25 Jul 2017 16:29:33	U	SMPL
I4400-03 - 1	SB02A	2783	0.7125	-	7471A	PPB	25 Jul 2017 16:31:51	U	SMPL
I4400-04 - 1	SB02B	194	0.0347	-	7471A	PPB	25 Jul 2017 16:34:09	U	SMPL
I4402-01 - 1	ROLL-OFFS-COMP	160	0.0258	-	7471A	PPB	25 Jul 2017 16:36:25	U	SMPL
CCV02 - 1	CCV02	19697	5.1407	-	7471A	PPB	25 Jul 2017 16:38:41	U	SMPL
CCB02 - 1	CCB02	-119	-0.0472	-	7471A	PPB	25 Jul 2017 16:40:57	U	SMPL
I4407-01 - 1	HDD-PILWC3	43	-0.0048	-	7471A	PPB	25 Jul 2017 16:43:13	U	SMPL
I4414-01 - 1	OR-03-072417	1403	0.3513	-	7471A	PPB	25 Jul 2017 16:45:29	U	SMPL
I4415-01 - 1	SU-02-072417	693	0.1654	-	7471A	PPB	25 Jul 2017 16:47:45	U	SMPL
I4417-01 - 1	FK-BPM-01-A3-A4-A5-A6	672	0.1599	-	7471A	PPB	25 Jul 2017 16:50:02	U	SMPL
I4418-01 - 1	FK-BPM-01-B3-B4-B5-B6	1408	0.3526	-	7471A	PPB	25 Jul 2017 16:52:19	U	SMPL
I4419-01 - 1	FK-BPM-01-A7-A8-B7-B8	1333	0.3329	-	7471A	PPB	25 Jul 2017 16:54:36	U	SMPL
I4419-01DUP - 1	FK-BPM-01-A7-A8-B7-B8DUP	1291	0.3219	-	7471A	PPB	25 Jul 2017 16:56:53	U	SMPL
I4419-01MS - 1	FK-BPM-01-A7-A8-B7-B8MS	15993	4.1709	-	7471A	PPB	25 Jul 2017 16:59:10	U	SMPL
I4419-01MSD - 1	FK-BPM-01-A7-A8-B7-B8MSD	16349	4.2641	-	7471A	PPB	25 Jul 2017 17:01:29	U	SMPL
I4420-06 - 1	2A-1-2-3-4-5-COMP-(0-10)	638	0.1510	-	7471A	PPB	25 Jul 2017 17:03:45	U	SMPL
CCV03 - 1	CCV03	19900	5.1938	-	7471A	PPB	25 Jul 2017 17:06:01	U	SMPL
CCB03 - 1	CCB03	-119	-0.0472	-	7471A	PPB	25 Jul 2017 17:08:17	U	SMPL
I4420-13 - 1	2A-6-7-8-9-10-COMP-(0-10)	646	0.1531	-	7471A	PPB	25 Jul 2017 17:10:34	U	SMPL
I4423-01 - 1	NR-P-1	14527	3.7871	-	7471A	PPB	25 Jul 2017 17:12:51	U	SMPL
I4423-02 - 1	NR-P-2	2862	0.7332	-	7471A	PPB	25 Jul 2017 17:15:07	U	SMPL
I4423-03 - 1	NR-P-3	4467	1.1534	-	7471A	PPB	25 Jul 2017 17:17:23	U	SMPL
I4423-04 - 1	NR-P-4	601	0.1413	-	7471A	PPB	25 Jul 2017 17:19:40	U	SMPL
PB100929BL - 1	PBS	4	-0.0150	-	7471A	PPB	25 Jul 2017 17:21:57	U	SMPL
PB100929BS - 1	LCSS	14411	3.7568	-	7471A	PPB	25 Jul 2017 17:24:14	U	SMPL
I4421-07 - 1	E2-HS5-NSW	26858	7.0154	-	7471A	PPB	25 Jul 2017 17:26:32	U	SMPL
I4421-08 - 1	E2-HS5-ESW	63793	16.6851	-	7471A	PPB	25 Jul 2017 17:28:50	U	SMPL
I4421-09 - 1	E2-HS5-SSW	14071	3.6678	-	7471A	PPB	25 Jul 2017 17:31:07	U	SMPL
CCV04 - 1	CCV04	19534	5.0980	-	7471A	PPB	25 Jul 2017 17:33:26	U	SMPL
CCB04 - 1	CCB04	-147	-0.0545	-	7471A	PPB	25 Jul 2017 17:35:42	U	SMPL
I4421-10 - 1	E2-HS5-WSW	25746	6.7243	-	7471A	PPB	25 Jul 2017 17:37:59	U	SMPL
I4421-11 - 1	E2-HS5-BASE	53433	13.9728	-	7471A	PPB	25 Jul 2017 17:40:16	U	SMPL
I4421-12 - 1	E2-M6-ESW	216387	56.6344	-	7471A	PPB	25 Jul 2017 17:42:33	U	SMPL
I4421-13 - 1	E2-N7-SSW	51687	13.5157	-	7471A	PPB	25 Jul 2017 17:44:52	U	SMPL
I4421-14 - 1	E2-M6-WSW	113601	29.7249	-	7471A	PPB	25 Jul 2017 17:47:22	U	SMPL
I4421-14DUP - 1	E2-M6-WSWDUP	111699	29.2269	-	7471A	PPB	25 Jul 2017 17:52:45	U	SMPL
I4421-14MS - 1	E2-M6-WSWMS	131528	34.4182	-	7471A	PPB	25 Jul 2017 18:03:20	U	SMPL
I4421-14MSD - 1	E2-M6-WSWMSD	124864	32.6735	-	7471A	PPB	25 Jul 2017 18:05:40	U	SMPL
I4422-05 - 1	VNJ-201-COMP	3221	0.8272	-	7471A	PPB	25 Jul 2017 18:08:01	U	SMPL
I4422-09 - 1	72-12018-COMP	1279	0.3188	-	7471A	PPB	25 Jul 2017 18:10:22	U	SMPL
CCV05 - 1	CCV05	18569	4.8453	-	7471A	PPB	25 Jul 2017 18:12:39	U	SMPL
CCB05 - 1	CCB05	-163	-0.0587	-	7471A	PPB	25 Jul 2017 18:14:55	U	SMPL
I4422-13 - 1	72-11940-COMP	761	0.1832	-	7471A	PPB	25 Jul 2017 18:17:12	U	SMPL
I4422-17 - 1	RO-156-COMP	2210	0.5625	-	7471A	PPB	25 Jul 2017 18:19:28	U	SMPL
I4424-01 - 1	AU-06-072417-A	699	0.1669	-	7471A	PPB	25 Jul 2017 18:21:45	U	SMPL
I4424-04 - 1	AU-06-072417-B	1026	0.2526	-	7471A	PPB	25 Jul 2017 18:24:01	U	SMPL
I4424-07 - 1	AU-06-072417-C	1329	0.3319	-	7471A	PPB	25 Jul 2017 18:26:18	U	SMPL
I4424-10 - 1	AU-06-072417-D	750	0.1803	-	7471A	PPB	25 Jul 2017 18:28:35	U	SMPL
I4419-01LX5 - 1		275	0.0559	-	7471A	PPB	25 Jul 2017 18:30:53	U	SMPL
I4419-01A - 1		17540	4.5759	-	7471A	PPB	25 Jul 2017 18:33:10	U	SMPL
I4421-14LX5 - 1		22211	5.7988	-	7471A	PPB	25 Jul 2017 18:35:28	U	SMPL
I4421-14A - 1		126009	32.9733	-	7471A	PPB	25 Jul 2017 18:37:44	U	SMPL
CCV06 - 1	CCV06	18973	4.9511	-	7471A	PPB	25 Jul 2017 18:40:01	U	SMPL
CCB06 - 1	CCB06	-209	-0.0708	-	7471A	PPB	25 Jul 2017 18:42:21	U	SMPL
I4421-08DLX5 - 1		12529	3.2641	-	7471A	PPB	25 Jul 2017 18:44:39	U	SMPL
I4421-11DLX5 - 1		10820	2.8166	-	7471A	PPB	25 Jul 2017 18:46:56	U	SMPL
I4421-12DLX10 - 1		23052	6.0190	-	7471A	PPB	25 Jul 2017 18:49:13	U	SMPL
I4421-13DLX5 - 1		10316	2.6847	-	7471A	PPB	25 Jul 2017 18:51:30	U	SMPL
I4421-14DLX5 - 1		22879	5.9737	-	7471A	PPB	25 Jul 2017 18:53:47	U	SMPL
I4421-14DUPDLX5 - 1		22691	5.9245	-	7471A	PPB	25 Jul 2017 18:56:04	U	SMPL
I4421-14MSDLX5 - 1		26600	6.9479	-	7471A	PPB	25 Jul 2017 18:58:22	U	SMPL
I4421-14MSDDLX5 - 1		25429	6.6413	-	7471A	PPB	25 Jul 2017 19:00:40	U	SMPL
I4421-14LDLX25 - 1		4314	1.1134	-	7471A	PPB	25 Jul 2017 19:02:59	U	SMPL
I4421-14ADLX5 - 1		26074	6.8102	-	7471A	PPB	25 Jul 2017 19:05:16	U	SMPL
CCV07 - 1	CCV07	18860	4.9215	-	7471A	PPB	25 Jul 2017 19:07:33	U	SMPL
CCB07 - 1	CCB07	-128	-0.0496	-	7471A	PPB	25 Jul 2017 19:09:51	U	SMPL